

This mail sent on 27th April 2017 from president_cmai@cmai.asia

Subject: Voting on IEC TC /86, 86B Documents (LITD 11) Optical fibre documents

Dear Sir,

we invite your views on enclosed documents pertaining to optical fiber equipment/devices etc. for BIS within next 2 days.

REGARDS

Prof N K Goyal

President CMAI Association of India, Chairman Emeritus TEMA

Member Governing Council Telecom Equipment & Services Export Promotion Council, Govt. of India

Vice Chairman, ITU APT India, Adviser Gujarat Technological University

Chairman India Trade Promotion Services, Dubai UAE,

Director NFL National Fertilizers Ltd, Govt. of India

+91 98 111 29879, president_cmai@cmai.asia, www.cmai.asia

----- Forwarded Message -----

From: jambholkar.bis <jambholkar@bis.gov.in>

Cc: ltd@bis.gov.in; ltd11@bis.gov.in

Sent: Tuesday, 25 April 2017 3:47 PM

Subject: Voting on IEC TC /86, 86B Documents (LITD 11)

To,

All Members of Fibre Optics, Fibres, Cables and Devices Sectional Committee, LITD 11

Dear Madam/Sir,

Please find enclosed herewith the following document(s) of IEC Committee and Sub Committees **86 Fibre optics, 86 Fibres and cables, and 86B: Fibre optic interconnecting devices and passive components** due for voting by the due date indicated against each document:

Sl No.	Doc./Ref. No.	Stage	Title	Due Date
1	86/511/CD	CD	Calibration of fibre-optic power meters	12/05/2017
2	86/514/CD	CD	Calibration of wavelength/optical frequency measurement instruments - Part 3:Optical frequency meters internally referenced to a frequency comb	30/06/2017
3	86B/4076/FDIS	FDIS	Fibre optic interconnecting devices and passive components - Performance standard - Part 121-2 : Simplex and duplex cords with single-mode fibre and cylindrical ferrule connectors for category C -	05/05/2017

2. India is a Participating (P) Member on IEC/ TC 86 and Observer (O) Member of 86B the Committee responsible for preparation of this(these) document(s). As P Member of IEC/TC 86, it is our obligation to vote on these documents. We can also vote/comment on documents even if we are O-members. LITD 11 acts as the national mirror committee of TC 86, 86A and 86B. You, being member of LITD 11, are requested to kindly examine this(these) document(s) and let us have your comments in the attached format, if any, for forwarding to TC 86, 86A and 86B.

3. You, being the member of LITD 11, are requested to kindly let us know whether we should vote in favour or against the document(s) or abstain from voting. Kindly send your reply within a week and comments if any in the attached format so as to enable us to send the India's viewpoint to IEC.

धन्यवाद/With warm regards,

बिपिन जांभोलकर/ Bipin Jambholkar

भारतीय मानक ब्यूरो/Bureau of Indian Standards

मानक भवन/ Manak Bhavan

9 बहादुरशाह जफर मार्ग/ 9 Bahadur Shah Zafar Marg

नई दिल्ली-110002/New Delhi - 110002

भारत/ INDIA

टेलीफाक्स Telefax - 91-11-23237093

फ़ोन /Phone - 91-11-23230131/विस्तारExtn: 4235,4401



86/511/CD

COMMITTEE DRAFT (CD)

PROJECT NUMBER: IEC 61315 ED3	
DATE OF CIRCULATION: 2017-01-20	CLOSING DATE FOR COMMENTS: 2017-05-12
SUPERSEDES DOCUMENTS: 86/492/CD,86/503A/CC	

IEC TC 86 : FIBRE OPTICS	
SECRETARIAT: United States of America	SECRETARY: Mr Steve Swanson
OF INTEREST TO THE FOLLOWING COMMITTEES:	PROPOSED HORIZONTAL STANDARD: <input type="checkbox"/> Other TC/SCs are requested to indicate their interest, if any, in this CD to the secretary.
FUNCTIONS CONCERNED: <input type="checkbox"/> EMC <input type="checkbox"/> ENVIRONMENT <input type="checkbox"/> QUALITY ASSURANCE <input type="checkbox"/> SAFETY	

This document is still under study and subject to change. It should not be used for reference purposes.

Recipients of this document are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

TITLE:
Calibration of fibre-optic power meters

NOTE FROM TC/SC OFFICERS:

CONTENTS

2		
3	FOREWORD.....	4
4	INTRODUCTION.....	6
5	1 Scope.....	7
6	2 Normative references.....	7
7	3 Terms and definitions	7
8	4 Preparation for calibration.....	15
9	4.1 Organization	15
10	4.2 Traceability	15
11	4.3 Advice for measurements and calibrations.....	15
12	4.4 Recommendations to users	16
13	5 Absolute power calibration	16
14	5.1 Calibration methods	16
15	5.2 Establishing the calibration conditions	17
16	5.3 Calibration procedure	18
17	5.4 Calibration uncertainty.....	19
18	5.4.1 Uncertainty due to the setup.....	19
19	5.4.2 Uncertainty of the reference meter	20
20	5.4.3 Correction factors and uncertainty caused by the change of conditions	21
21	5.4.4 Uncertainty due to the test meter.....	24
22	5.5 Reporting the results	25
23	6 Measurement uncertainty of a calibrated power meter	25
24	6.1 Overview.....	25
25	6.2 Uncertainty at reference conditions.....	25
26	6.3 Uncertainty at operating conditions.....	26
27	6.3.1 Determination of dependences on conditions.....	26
28	6.3.2 Ageing	27
29	6.3.3 Dependence on temperature	27
30	6.3.4 Dependence on the power level (nonlinearity).....	28
31	6.3.5 Dependence on the type of fibre or on the beam geometry	28
32	6.3.6 Dependence on the connector-adaptor combination	30
33	6.3.7 Dependence on wavelength	30
34	6.3.8 Dependence on spectral bandwidth	31
35	6.3.9 Dependence on polarization	31
36	6.3.10 Other dependences.....	32
37	7 Nonlinearity calibration.....	32
38	7.1 General.....	32
39	7.2 Nonlinearity calibration based on superposition.....	33
40	7.2.1 Procedure.....	33
41	7.2.2 Uncertainties	34
42	7.3 Nonlinearity calibration based on comparison with a calibrated power meter	35
43	7.3.1 Procedure.....	35
44	7.3.2 Uncertainties	36
45	7.4 Nonlinearity calibration based on comparison with an attenuator	36
46	7.5 Calibration of power meter for high power measurement.....	36
47	Annex A (normative) Mathematical basis for measurement uncertainty calculations	37
48	A.1 General.....	37

49	A.2	Type A evaluation of uncertainty.....	37
50	A.3	Type B evaluation of uncertainty.....	38
51	A.4	Determining the combined standard uncertainty.....	38
52	A.5	Reporting	39
53	Annex B (informative)	Linear to dB scale conversion of uncertainties	40
54	B.1	Definition of decibel.....	40
55	B.2	Conversion of relative uncertainties	40
56	Bibliography		41
57			
58	Figure 1	– Typical spectral responsivity of photoelectric detectors.....	13
59	Figure 2	– Example of a traceability chain.....	14
60	Figure 3	– Measurement setup for sequential, fibre-based calibration	17
61	Figure 4	– Change of conditions and uncertainty.....	21
62	Figure 5	– Determining and recording an extension uncertainty.....	27
63	Figure 6	– Possible subdivision of the optical reference plane into 10 x 10 squares, for 64 the measurement of the spatial response	28
65	Figure 7	– Wavelength dependence of response due to Fabry-Perot type interference	31
66	Figure 8	– Measurement setup of polarization dependent response	32
67	Figure 9	– Nonlinearity calibration based on superposition	33
68	Figure 10	– Measurement setup for nonlinearity calibration by comparison.....	35
69			
70	Table 1	– Typical calibration methods and correspondent power	16
71	Table 2	– Nonlinearity	34
72			

INTERNATIONAL ELECTROTECHNICAL COMMISSION

CALIBRATION OF FIBRE-OPTIC POWER METERS
FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with an IEC Publication.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.
- International Standard IEC 61315 has been prepared by IEC technical committee 86: Fibre optics.
- This third edition cancels and replaces the second edition published in 2005. It constitutes a technical revision.
- This edition includes the following significant technical changes with respect to the previous edition:
- update of terms and definitions;
 - update of Section 5.1 including Table 1 (new type of source);
 - update of Annex A;
 - addition of Annex B on dB conversion.

The text of this standard is based on the following documents:

FDIS	Report on voting
86/xxx/FDIS	86/xxx/RVD

124
125 Full information on the voting for the approval of this standard can be found in the report on
126 voting indicated in the above table.

127 This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

128 The committee has decided that the contents of this publication will remain unchanged until the
129 maintenance result date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data
130 related to the specific publication. At this date, the publication will be

- 131 • reconfirmed,
- 132 • withdrawn,
- 133 • replaced by a revised edition, or
- 134 • amended.

135

136

137

INTRODUCTION

138 Fibre-optic power meters are designed to measure optical power from fibre-optic sources as
139 accurately as possible. This capability depends largely on the quality of the calibration process.
140 In contrast to other types of measuring equipment, the measurement results of fibre-optic
141 power meters usually depend on many conditions of measurement. The conditions of
142 measurement during the calibration process are called *calibration conditions*. Their precise
143 description shall therefore be an integral part of the calibration.

144 This International Standard defines all of the steps involved in the calibration process:
145 establishing the *calibration conditions*, carrying out the calibration, calculating the uncertainty,
146 and reporting the uncertainty, the *calibration conditions* and the *traceability*.

147 The absolute power calibration describes how to determine the ratio between the value of the
148 input power and the power meter's result. This ratio is called *correction factor*. The
149 measurement uncertainty of the correction factor is combined following Annex A from
150 uncertainty contributions from the reference meter, the test meter, the setup and the
151 procedure.

152 The calculations go through detailed characterizations of individual uncertainties. It is important
153 to know that:

- 154 a) experience-based, type B estimations of the individual uncertainties are acceptable;
- 155 b) a detailed uncertainty analysis is only necessary once for each power meter type under
156 test, and all subsequent calibrations can be based on this one-time analysis, using the
157 appropriate type A measurement contributions evaluated at the time of the calibration;
- 158 c) some of the individual uncertainties can simply be considered to be part of a checklist, with
159 an actual value which can be neglected.

160 Calibration according to Clause 5 is mandatory for reports referring to this standard.

161 Clause 6 describes the evaluation of the measurement uncertainty of a calibrated power meter
162 operated within *reference conditions* or within *operating conditions*. It depends on the
163 calibration uncertainty of the power meter as calculated in 5.4, the conditions and its
164 dependence on the conditions. It is usually performed by manufacturers in order to establish
165 specifications and is not mandatory for reports referring to this standard. One of these
166 dependences, the nonlinearity, is determined in a separate calibration (Clause 7).

167

CALIBRATION OF FIBRE-OPTIC POWER METERS

168
169
170

171 **1 Scope**

172 This international standard is applicable to instruments measuring radiant power emitted from
173 sources that are typical for the fibre-optic communications industry. These sources include
174 laser diodes, light emitting diodes (LEDs) and fibre-type sources. The radiation may be
175 divergent or collimated. The standard describes the calibration of power meters to be
176 performed by calibration laboratories or by power meter manufacturers.

177 **2 Normative references**

178 The following documents are referred to in the text in such a way that some or all of their
179 content constitutes requirements of this document. For dated references, only the edition cited
180 applies. For undated references, the latest edition of the referenced document (including any
181 amendments) applies.

182 IEC 60359:2001, *Electrical and electronic measurement equipment – Expression of*
183 *performance*

184 IEC 60793-2, *Optical fibres – Part 2: Product specifications – General*

185 IEC TR 61931, *Fibre optic – Terminology*

186 ISO/IEC 17025:2005, *General requirements for the competence of testing and calibration*
187 *laboratories*

188 VIM: JCGM 200:2012, *International vocabulary of metrology – Basic and general concepts and*
189 *associated terms*

190 GUM: JCGM 100:2008, *Evaluation of measurement data – Guide to the expression of*
191 *uncertainty in measurement*

192 **3 Terms and definitions**

193 For the purposes of this document, the terms and definitions given in IEC TR 61931 and the
194 following apply.

195 ISO and IEC maintain terminological databases for use in standardization at the following
196 addresses:

- 197 • IEC Electropedia: available at <http://www.electropedia.org/>
- 198 • ISO Online browsing platform: available at <http://www.iso.org/obp>

199

200 **3.1**

201 **accredited calibration laboratory**

202 calibration laboratory authorized by the appropriate national organization to issue calibration
203 certificates with a minimum specified uncertainty, which demonstrate traceability to national
204 standards (3.14)

205 **3.2**
 206 **adjustment**
 207 set of operations carried out on an instrument in order that it provides given indications
 208 corresponding to given values of the measurand

209 Note to entry: When the instrument is made to give a null indication corresponding to a null value of the
 210 measurand, the set of operations is called zero adjustment.

211 [IEV 311-03-16; see also VIM 3.11, modified]

212 **3.3**
 213 **calibration**
 214 set of operations that establish, under specified conditions, the relationship between the values
 215 of quantities indicated by a measuring instrument and the corresponding values realized by
 216 standards

217 Note 1 to entry: The result of a calibration permits either the assignment of values of measurands to the indications
 218 or the determination of corrections with respect to indications.

219 Note 2 to entry: A calibration may also determine other metrological properties such as the effect of influence
 220 quantities.

221 Note 3 to entry: The result of a calibration may be recorded in a document, sometimes called a calibration
 222 certificate or a calibration report.

223 [See also VIM 2.39, modified]

224 **3.4**
 225 **calibration conditions**
 226 conditions of measurement in which the calibration is performed

227 **3.5**
 228 **centroidal wavelength**
 229 λ_c
 230 power-weighted mean wavelength of a light source in vacuum.

231 Note 1 to entry: For a continuous spectrum, the centroidal wavelength is defined as:

$$232 \quad \lambda_c = \frac{\int p(\lambda)\lambda d\lambda}{P_{\text{total}}} \quad (1)$$

233 For a spectrum consisting of discrete lines, the centroidal wavelength is defined as:

$$234 \quad \lambda_c = \frac{\sum_i P_i \lambda_i}{\sum_i P_i} \quad (2)$$

235 where

236 $p(\lambda)$ is the power spectral density of the source, for example, in W/nm;

237 λ_i is the vacuum wavelength of the i^{th} discrete line;

238 P_i is the power of the i^{th} discrete line, for example, in W;

239 P_{total} is the total power, for example, in W.

240 Note 2 to entry: The above integrals and summations theoretically extend over the entire spectrum of the light
 241 source. However, it is usually sufficient to perform the integral or summation over the spectrum where the spectral
 242 density $p(\lambda)$ or power P_i is higher than 0,1 % of the maximum spectral density $p(\lambda)$ or power P_i .

243 **3.6**
 244 **correction factor**
 245 *CF*
 246 numerical factor by which the uncorrected result of a measurement is multiplied to compensate
 247 for systematic error

248 [See also VIM]

249 **3.7**
 250 **detector**
 251 element of the power meter that transduces the radiant optical power into a measurable,
 252 usually electrical, quantity. In this standard, the detector is assumed to be connected with the
 253 optical input port by an optical path

254 [SOURCE: IEC 61931; see also VIM, 3.9, modified]

255 **3.8**
 256 **deviation**
 257 *D*
 258 for the purposes of this standard, the relative difference between the power measured by the
 259 test meter (3.32) P_{DUT} and the reference power P_{ref}

$$260 \quad D = \frac{P_{\text{DUT}} - P_{\text{ref}}}{P_{\text{ref}}} \quad (3)$$

261 **3.9**
 262 **excitation (fibre)**
 263 description of the distribution of optical power between the modes in the fibre. In context with
 264 multimode fibres, the fibre excitation is described by:

- 265 a) the spot diameter (3.31) on the surface of the fibre end, and
- 266 b) the numerical aperture (3.17) of the radiation emitted from the fibre.

267 Single mode fibres are generally assumed to be excited by only one mode (the fundamental
 268 mode)

269 **3.10**
 270 **instrument state**
 271 set of parameters that can be chosen on an instrument

272 Note to entry: Typical parameters of the instrument state are the optical power range, the wavelength setting, the
 273 display measurement unit and the output from which the measurement result is obtained (for example display,
 274 interface bus, analogue output).

275 **3.11**
 276 **irradiance**
 277 quotient of the incremental radiant power ∂P incident on an element of the reference plane by
 278 the incremental area ∂A of that element:

$$279 \quad E = \frac{\partial P}{\partial A} \quad (\text{W/m}^2) \quad (4)$$

280 [SOURCE: IEC TR 61931, definition 2.1.15, mainly modified by adding the equation]

281 **3.12**282 **measurement result**283 *y*

284 (displayed or electrical) output of a power meter (or standard), after completing all actions
 285 suggested by the operating instructions, for example warm-up, zero adjustment and
 286 wavelength-correction

287 Note to entry: Measurement result is expressed in watts (W). For the purposes of uncertainty, measurement results
 288 in other units, for example volts, should be converted to watts. Measurement results in decibels (dB) should also be
 289 converted to watts, because the entire uncertainty accumulation is based on measurement results expressed in
 290 watts.

291 **3.13**292 **measuring range**

293 set of values of measurands for which the error of a measuring instrument is intended to lie
 294 within specified limits

295 Note to entry: In this standard, the measuring range is the range of radiant power (part of the *operating range*), for
 296 which the uncertainty at operating conditions is specified. The term "dynamic range" should be avoided in this
 297 context.

298 [See also VIM 4.7, modified]

299 **3.14**300 **national (measurement) standard**

301 standard recognized by a national decision to serve, in a country, as the basis for assigning
 302 values to other standards of the quantity concerned

303 [See also VIM 5.3, modified]

304 **3.15**305 **national standards laboratory**

306 laboratory which maintains the national standard (3.14)

307 **3.16**308 **nonlinearity**309 *NL*

310 relative difference between the response (3.28) at a given power *P* and the response at a
 311 reference power *P*₀:

$$312 \quad nI_{P/P_0} = \frac{r(P)}{r(P_0)} - 1 \quad (5)$$

313 If expressed in decibels, the nonlinearity is:

$$314 \quad NL_{P/P_0} = 10 \times \log_{10} \frac{r(P)}{r(P_0)} \quad (\text{dB}) \quad (6)$$

315 Note 1 to entry: The nonlinearity is equal to zero at the reference power.

316 Note 2 to entry: The term "local nonlinearity" is used for the relative difference between the responses at two
 317 different power levels (separated by 3,01 dB) obtained during the nonlinearity calibration. The term "global
 318 nonlinearity" is used for the result of summing the local nonlinearities (in dB); it is identical to the nonlinearity
 319 defined here.

320 **3.17**321 **numerical aperture**

322 description of the beam divergence of an optical source. In this standard, the numerical
323 aperture is the sine of the (linear) half-angle at which the irradiance is 5 % of the maximum
324 irradiance.

325 [SOURCE: IEC 60793-1-43, modified – adapted from the definition of the numerical aperture of
326 multimode graded-index fibres in IEC 60793-1-43; in this standard, the definition is used to
327 describe the divergence of all divergent beams]

328 **3.18**329 **operating conditions**

330 appropriate set of specified ranges of values of influence quantities usually wider than the
331 reference conditions for which the uncertainties of a measuring instrument are specified

332 Note to entry: The operating conditions and uncertainty at operating conditions are usually specified by
333 manufacturer for the convenience of the user.

334 **3.19**335 **operating range**

336 specified range of values of one of a set of operating conditions (3.18)

337 **3.20**338 **optical input port**

339 physical input of the power meter (or standard) to which the radiant power is to be applied or to
340 which the optical fibre end is to be connected. An optical path (path of rays with or without
341 optical elements like lenses, diaphragms, light guides, etc.) is assumed to connect the optical
342 input port with the power meter's detector.

343 **3.21**344 **optical reference plane**

345 plane on or near the optical input port (3.20) which is used to define the beam's spot diameter
346 (3.31)

347 Note to entry: The optical reference plane is usually assumed to be perpendicular to the beam propagation, and it
348 should be described by appropriate mechanical dimensions relative to the power meter's optical input port.

349 **3.22**350 **polarization dependent response**351 ***PDR***

352 variation in response of a power meter with respect to all possible polarization states of the
353 input light

$$354 \quad PDR = 10 \times \log_{10} \left(\frac{r_{\max}}{r_{\min}} \right) \quad (\text{dB}) \quad (7)$$

355 where r_{\max} and r_{\min} are the maximum and minimum response (3.28) taken over all polarization
356 states.

357 Note to entry: Polarization dependent response is expressed in decibels.

358 **3.23**359 **power meter (fibre-optic)**

360 instrument capable of measuring radiant power from fibre-coupled sources such as lasers and
361 LEDs, which are typical for the fibre-optic communications industry. The radiation may be
362 divergent or collimated. The radiation is assumed to be incident on the optical reference plane
363 within the specified conditions. A power meter may consist of either a single instrument or a
364 main instrument and a separate sensing head. In the case of a separate sensing head, the
365 head may be calibrated without the main instrument.

366 Note 1 to entry: The measurement result may be influenced by the main instrument, particularly if any analog
 367 electronics is used in the main instrument. In such cases, the sensing head shall be calibrated together with the
 368 main instrument.

369 Note 2 to entry: A fibre-optic power meter is usually capable of measuring the time-average of modulated optical
 370 power. An increased uncertainty may be observed, which depends on the duty cycle and the peak power of
 371 modulated optical power.

372 **3.24**
 373 **radiant power**

374 ***P***
 375 power emitted, transferred, or received in the form of optical radiation [1]¹⁾.

376 Note to entry: Radiant power is expressed in watts.

377 **3.25**
 378 **reference conditions**

379 conditions of use prescribed for testing the performance of a measuring instrument or for
 380 intercomparison of results of measurements

381 Note to entry: The reference conditions generally include reference values or reference ranges for the influence
 382 quantities affecting the measuring instrument.

383 **3.26**
 384 **reference meter**

385 standard which is used as the reference for the calibration (3.3) of a test meter (3.32)

386 **3.27**
 387 **reference standard**

388 standard, generally having the highest metrological quality available at a given location or in a
 389 given organization, from which measurements made there are derived

390 [See also VIM 5.6, modified]

391 **3.28**
 392 **response**

393 ***r***
 394 measurement result of a power meter, *y*, divided by the radiant power on the power meter's
 395 optical reference plane, *P*, at a given condition of measurement:

$$396 \quad r = \frac{y}{P} \quad (\text{W/W, dimensionless}) \quad (8)$$

397 Note to entry: An ideal power meter exhibits a response of 1 for all operating conditions.


398 **3.29**
 399 **(spectral) responsivity**

400 ***R***
 401 quotient of the detector output current *I* by the incident monochromatic optical power *P*:

$$402 \quad R = \frac{I}{P} \quad (\text{A/W}) \quad (9)$$

403 Note to entry: The responsivity depends on the conditions (wavelength, temperature, etc.).

1) Figures in square brackets refer to the Bibliography.



404

IEC 1826/05

405 **Key**

406 Si: silicon

407 Ge: germanium

408 InGaAs: indium gallium arsenide

409 **Figure 1 – Typical spectral responsivity of photoelectric detectors**410 **3.30**411 **spectral bandwidth**412 ***B***

413 full-width at half-maximum (FWHM) of the source spectrum

414 Note 1 to entry: If the source is a laser diode with a multiple-longitudinal mode spectrum, then the FWHM spectral
 415 bandwidth *B* is the rms spectral bandwidth, multiplied by 2,35 (assuming the source has a Gaussian envelope).

$$416 \quad B = 2,35 \sqrt{\frac{1}{P_{\text{total}}} \sum_i P_i (\lambda_i - \lambda_c)^2} \quad (10)$$

$$417 \quad P_{\text{total}} = \sum_i P_i \quad (11)$$

418 where

419 λ_c is the centroidal wavelength (3.5) of the laser diode, in nm;420 P_{total} is the total power, in W;421 P_i is the power of i^{th} longitudinal mode, in W;422 λ_i is the wavelength of i^{th} longitudinal mode, in nm.

423 Note 2 to entry: If the source emits at one wavelength only (single-line spectrum), it may be sufficient to specify an
 424 upper limit, for example spectral bandwidth < 1 nm.

425 Note 3 to entry: It is usually sufficient to perform the integral or summation over the spectrum where the power is
 426 higher than 0,1 % of the maximum power.

427 **3.31**428 **spot diameter**

429 in this standard, diameter of the irradiated area on the optical reference plane, defined by the
 430 (best-approximation) circle at which the irradiance (3.11) has dropped to 5 % of the peak
 431 irradiance

432 Note 1 to entry: The ratio of 5 % was adopted for reasons of compatibility with the definition of the numerical
 433 aperture. Other ratios are often used to describe laser beams, for example $1/e^2$ or $1/e$. In that case, it shall be
 434 stated with the spot diameter value.


435 **3.32**
 436 **test meter**
 437 power meter (3.23) (or standard) to be calibrated by comparison with the reference meter
 438 (3.26)

439 **3.33**
 440 **traceability**
 441 property of the result of a measurement or the value of a standard whereby it can be related to
 442 stated references, usually national or international standards, through an unbroken chain of
 443 comparisons all having stated uncertainties

444 [See also VIM 2.41, modified]

445 **3.34**
 446 **traceability chain**
 447 unbroken chain of comparison

448 [See also VIM 2.42, modified]



IEC 1827/05

449

450 **Figure 2 – Example of a traceability chain**

451 **3.35**
 452 **working standard**
 453 standard that is used routinely to calibrate or check measuring instruments

454 [See also VIM 5.7, modified]

455 Note to entry: A working standard is usually calibrated against a reference standard (3.27).

456 **3.36**
 457 **zero error**
 458 measurement result of a power meter without irradiation of the optical input port

459 [See also VIM 4.28, modified]

460 **4 Preparation for calibration**

461 **4.1 Organization**

462 The calibration laboratory should satisfy requirements of ISO/IEC 17025.

463 There should be a documented measurement procedure for each type of calibration performed,
464 giving step-by-step operating instructions and equipment to be used.

465 **4.2 Traceability**

466 The requirements of ISO/IEC 17025 should be met.

467 All standards used in the calibration process shall be calibrated according to a documented
468 program with *traceability to national standards laboratories* or to *accredited calibration*
469 *laboratories*. It is advisable to maintain more than one standard on each hierarchical level, so
470 that the performance of the standard can be verified by comparisons on the same level. Make
471 sure that any other test equipment which has a significant influence on the calibration results is
472 calibrated. Upon request, specify this test equipment and its *traceability chain(s)*. The re-
473 calibration period(s) shall be defined and documented.

474 **4.3 Advice for measurements and calibrations**

475 This subclause gives general advice for all measurements and calibrations of optical and fibre-
476 optic power meters.

477 The calibration should be made in a temperature-controlled room if non-temperature-controlled
478 detectors are used. The recommended temperature is 23 °C. Humidity control may be
479 necessary if humidity-sensitive optical detectors are used, or if there is the possibility of
480 condensation on the components. A change of the laboratory's humidity may change the
481 absorption of air and thereby change the power. This effect is relatively strong between 1 360 nm
482 and 1 410 nm, especially when a sequential-type, open-beam calibration is used and the humidity
483 changes between the steps. In parallel-type calibrations with open-beam paths of approximately
484 the same lengths, the measurement results of both the reference meter and the test meter will
485 change at approximately the same time, with negligible effect on the calibration result.

486 The laboratory should be kept clean. Connectors and optical input ports should always be
487 cleaned before measurement. The quality and cleanness of the connector in front of the
488 detector should be checked. All fibres should be moved as little as possible during the
489 measurements; they can be fixed to the workbench if necessary. Sensors should be moved to
490 the fibre rather than the fibre to the sensor.

491 The optical source that is used for the excitation of the power meter should be characterized
492 for centroidal wavelength and spectral bandwidth. The spectral bandwidth should be narrow
493 enough to avoid averaging over a wide range of wavelengths. Means to ensure the stability of
494 the source, for example with the help of independent power monitoring, may be advisable.

495 Laser diodes are sensitive to back reflections. To improve the stability, it is advisable to use an
496 optical attenuator or an optical isolator between the laser diode and the test meter. Because of
497 their narrow spectral bandwidths, the combination of laser diode and multimode fibre is also
498 capable of producing speckle patterns on the optical reference plane, with the result of an
499 increased measurement uncertainty.

500 Fibre connectors and connector adapters are likely to produce errors in the measurement
501 result [2] because of multiple reflections between the *optical input port* (or detector) and the
502 connector-adapter combination (as part of the source). Therefore, connectors and adapters
503 with low reflectivity are recommended for the calibration. Otherwise, a correction factor and an
504 increased uncertainty may have to be taken into account.

505 It is advisable to use reference meters with detector diameters of ≥ 3 mm, because they can
 506 easily be irradiated with an open beam, and they are less susceptible to contamination (dirt and
 507 dust). The reference meter's surface reflections should be as small as possible. If the source
 508 emits a divergent beam, then a reference meter with an integrating sphere may be advisable. It
 509 is also acceptable to use meters with "flat" detectors and mathematical correction, based on
 510 multiplying the emitted far field distribution with the measured angle-dependence of the
 511 detector of the reference meter, and integrating over the range of far-field angles.

512 Temperature control of the detectors should be considered for highly accurate calibrations,
 513 because detectors exhibit strong temperature dependence over some wavelength ranges.

514 4.4 Recommendations to users

515 It is recommended that the user of the power meter maintain at least one reference power
 516 meter, which allows comparison of the meters for confidence. These comparisons are
 517 particularly important before and after the meter is sent to re-calibration, because they will
 518 allow the user to determine whether or not his scale has changed – for example due to
 519 transport – after the meter returns. Scale changes due to adjustment (3.2) (see IEV 311-03-16
 520 and VIM 4.30) will be reported on the calibration certificate.

521 A regular comparison of the correction factors (3.6), or of the deviations (3.8), will allow the
 522 user to screen out excessive ageing, and possibly to adjust the recalibration intervals.

523 5 Absolute power calibration

524 5.1 Calibration methods

525 The calibration of a power meter is usually achieved by exposing both the meter under test and
 526 a calibrated power meter with known uncertainty (the reference meter) to an optical radiation,
 527 and by transferring the reference meter's (3.26) measurement result to the test meter (3.32).

528 The allowable spectral bandwidth (3.30) depends on the test meter's spectral responsivity
 529 (3.29); the stronger its wavelength dependence, the narrower the spectral bandwidth. Usual
 530 bandwidths are ≤ 10 nm, which excludes the possibility of calibrating with wider-bandwidth
 531 LEDs. Therefore, either combinations of "white-light" sources and narrow-bandwidth filters (for
 532 example monochromators), laser diodes or combinations of supercontinuum lasers with
 533 tuneable bandpass filters are used in fibre-optic power meter calibrations.

534 Depending on the type of source and the exciting beam geometry, six most frequent calibration
 535 methods can be distinguished:

536 **Table 1 –Calibration methods and correspondent typical power**


Radiation source	Open-beam calibration	Fibre beam calibration
"White-light" with filter	$P \approx 10 \mu\text{W}$	$P \approx 10 \text{ nW}$ to $0,3 \mu\text{W}$ (MMF) $P \approx 2 \text{ nW}$ (SMF)
Laser diode	$P \approx 10 \mu\text{W}$ to a few mW	$P \approx 10 \mu\text{W}$ to a few mW (SMF and MMF)
Supercontinuum laser with filter	$P \approx 1 \mu\text{W}$ to a few mW	$P \approx 1 \mu\text{W}$ to $700 \mu\text{W}$ (SMF and MMF)
MMF: multimode fibre (usually graded-index fibre) SMF: single-mode fibre		

537

538 For fibre-optic power meters, fibre beam calibration is recommended. For open-beam
 539 calibration, a correction of the calibration results using a series of fibre beam laser calibration
 540 results at a few wavelengths should be performed.

541 One can distinguish between the sequential and the parallel measurement method. When
 542 reference meter and test meter are sequentially exposed to the source, then the radiated
 543 power should be kept as constant as possible, for example by appropriate stabilization. For the
 544 parallel-type calibration, a beam splitter or a branching device is used to generate two beams
 545 that excite both the reference meter and the test meter simultaneously. In this case, the beam
 546 splitter or branching device ratio shall be determined as accurately as possible, and its stability
 547 shall be investigated.

548 As an example, a measurement setup for sequential, fibre-based calibration is illustrated in
 549 Figure 3. A launching device, for removal of the cladding modes and creation of an appropriate
 550 modal excitation, is included in the setup.



IEC 1828/05 modified

551

552 **Figure 3 – Measurement setup for sequential, fibre-based calibration**

553 5.2 Establishing the calibration conditions

554 The calibration conditions (3.4) are the measurement conditions during the calibration process.
 555 Establishing and maintaining the calibration conditions is an important part of the calibration
 556 (3.3), because any change of these conditions is capable of producing erroneous measurement
 557 results. The calibration conditions should be a close approximation to the intended operating
 558 conditions (3.18). This ensures that the (additional) uncertainty in the operating environment is
 559 as small as possible. The calibration conditions should be specified in the form of nominal
 560 values with uncertainties when applicable. In order to meet the requirements of this standard,
 561 the calibration conditions shall at least consist of:

- 562 a) the date of calibration;
- 563 b) the ambient temperature with uncertainty, for example $23\text{ °C} \pm 1\text{ °C}$;
- 564 c) the ambient relative humidity, if it has an influence, otherwise a relative humidity below the
 565 condensation point is assumed;
- 566 d) the nominal radiant power on the optical reference plane (3.21);
- 567 e) the beam geometry:
 - 568 1) an open (for example collimated) beam, described by the spot diameter (3.31) on the
 569 optical reference plane, the beam's numerical aperture (3.17) and the irradiance (3.11)
 570 distribution in the beam. Typical irradiance distributions are: uniform, Gaussian or even
 571 irregular (speckled);
 - 572 2) the type of fibre and, if applicable, its degree of excitation (for example within encircled
 573 flux templates defined in IEC 61280-4-1 when using an A1a or A1b multimode fibre);
- 574 f) the connector-adapter combination: the connector type, polishing and adapter as part of the
 575 exciting source (if applicable);
- 576 g) the centroidal wavelength (3.5) of the exciting source with its uncertainty;

- 577 h) the spectral bandwidth (3.30) of the exciting source with its uncertainty;
 578 i) the state of polarization: "unpolarized light" or "polarized light, indefinite state". If the latter
 579 is chosen, the uncertainty due to polarization dependent response (3.22) shall be taken into
 580 account in 5.4.2 and 5.4.4.

581 The above conditions may not be exhaustive. There may be other parameters that have a
 582 significant influence on the calibration uncertainty and therefore shall be reported, too.

583 In the calibration with an **open-beam**, the power meter's *optical reference plane* (3.21) should
 584 be centrally irradiated with a beam diameter smaller than the active area of the optical
 585 reference plane.

586 In the calibration with a fibre, a single-mode fibre or a multimode fibre may be used. A single-
 587 mode fibre may be advantageous because of its reproducible beam characteristics, but may
 588 not be available for all wavelengths. If a multimode fibre is used, then excitation between 85 %
 589 and 95 % (slightly underfilled) is preferred because this excitation can be more easily
 590 reproduced (encircled flux templates defined in IEC 61280-4-1 are a good example of this
 591 condition). A launching device may be necessary to create the appropriate excitation. Note that
 592 multimode fibres will emit irregular beam patterns (speckle patterns) when driven by a laser
 593 diode; this will result in an increased calibration uncertainty.

594 A connector-adaptor combination should only be reported if the power meter is calibrated with
 595 a fibre, not with an open beam. It is recommended to use a combination of connector and
 596 adaptor with sufficiently low reflections back to the power meter.

597 **5.3 Calibration procedure**

- 598 (1) Establish and record the appropriate calibration conditions (5.2). Switch on all
 599 instrumentation and wait for enough time to stabilize.
- 600 (2) Set up the instrument state (3.10) of the reference meter and test meter according to the
 601 instruction manual. Set the wavelength on all instruments for the source wavelength. Select
 602 appropriate power ranges. Record the instrument states of both meters. Adjust the zero of
 603 both meters if applicable.
- 604 (3) Measure the optical power with the reference meter $P_{\text{std},1}$. Multiply the measurement result
 605 by the correction factor of the reference meter CF_{std} reported in its calibration certificate if it
 606 has not been adjusted. Multiply by the correction factor CF_{change} calculated in 5.4.3 if
 607 necessary. Record the measurement result, $P_{\text{ref},1} = P_{\text{std},1} \times CF_{\text{std}} \times CF_{\text{change}}$.
- 608 (4) Measure the optical power with the test meter. Apply necessary corrections as suggested
 609 by the operating instructions. Record the measurement result, $P_{\text{DUT},1}$.
- 610 (5) Calculate the first of a series of correction factors:

$$611 \quad CF_{\text{comparison},1} = \frac{P_{\text{ref},1}}{P_{\text{DUT},1}} \quad (12)$$

- 612 (6) Repeat steps (3) through (5) several times, with the result of obtaining several correction
 613 factors, $CF_{\text{comparison},1}$ to $CF_{\text{comparison},n}$.
- 614 (7) Calculate and record the average correction factor, CF_{DUT} from the individual correction
 615 factors:

$$616 \quad CF_{\text{DUT}} = \frac{1}{n} \cdot \sum_{i=1}^n CF_{\text{comparison},i} \quad (13)$$

617 If desired the deviation D can be calculated from the correction factor:

$$618 \quad D = \frac{1}{CF_{\text{DUT}}} - 1 \quad (14)$$

619 In later use of the test meter, the measurement results shall be multiplied with CF_{DUT} .
 620 Alternatively, an adjustment (3.2) of the test meter can be made so that the correction factor is
 621 changed to 1. In this case, the comparison should be repeated for verification.

622 **5.4 Calibration uncertainty**

623 The calibration uncertainty is the measurement uncertainty of the correction factor CF_{DUT} .
 624 Calculate the combined standard uncertainty from:

$$625 \quad u(CF_{\text{DUT}}) = \sqrt{u_{\text{setup}}^2 + u_{\text{ref}}^2 + u_{\text{DUT}}^2} \quad (15)$$

626 where

627 u_{setup} = uncertainty due to the setup, (5.4.1);

628 u_{ref} = uncertainty of the reference meter, (5.4.2);

629 u_{DUT} = uncertainty due to the test meter, (5.4.4).

630 NOTE Equation (15) is valid only if the input quantities are independent or uncorrelated. If some input quantities
 631 are significantly correlated, take the correlation into account. See GUM for more detail.

632 Then calculate the expanded uncertainty from:

$$633 \quad U(CF_{\text{DUT}}) = k \times u(CF_{\text{DUT}}) \quad (16)$$

634 where k is the coverage factor. See Annex A for more detail.

635 **5.4.1 Uncertainty due to the setup**

636 The following uncertainties may come from the setup:

- 637 a) Uncertainty due to the source power instability. In addition to the intrinsic variation of output
 638 power versus time, a laser source may react with unstable power to variations of back-
 639 reflections and variations of the state of polarisation of back-reflected light.
- 640 b) Uncertainty due to the beam splitter or branching device ratio (for parallel method), for
 641 example due to their polarization dependence.
- 642 c) Depending on the setup and method, other uncertainties may have to be taken into
 643 account.

644 Instability of the source power, of the beam splitter or branching device ratio (for parallel
 645 method) will cause a scatter in the measurement of the correction factor. The uncertainty due
 646 to these instabilities can be calculated from the experimental standard deviation of the
 647 correction factors $CF_{\text{comparison},1}$ to $CF_{\text{comparison},n}$ measured during the calibration (Equation (12)).
 648 The number of comparisons should be large to reduce this uncertainty. See Annex A for more
 649 detail on type A evaluation of uncertainty.

$$650 \quad u_{\text{setup,typeA}} = \frac{s(CF_{\text{comparison}})}{\sqrt{n}} \quad (17)$$

651 where:

652 $s(CF_{\text{comparison}})$ is the experimental standard deviation of the correction factors;

653 n is the number of measurement cycles during the calibration process.

654 This uncertainty can also be calculated from a standard deviation evaluated once from
655 measurements and used for all calibrations or from a type B evaluation. The instability should
656 therefore not vary too much from one calibration to the next and not depend on the test meter.
657 The number n in Equation (17) is always the number of measurement cycles during the current
658 calibration process.

659 This type A evaluated uncertainty will also be influenced by the repeatability of the connection
660 when using a sequential measurement method or by slight changes in the measurement
661 conditions during the calibration process. It can (partially) take into account some of the
662 uncertainties due to the reference meter (5.4.2) or test meter (5.4.4). Uncertainty components
663 should not be taken twice into account but also not be forgotten.

664 Calculate the uncertainty due to the setup by combining all partial uncertainties described in
665 this subclause:

$$666 \quad u_{\text{setup}} = \sqrt{\sum_{i=1}^m u_{\text{setup},i}^2} \quad (18)$$

667 **5.4.2 Uncertainty of the reference meter**

668 The uncertainty of the reference meter is mainly due to its calibration, to the uncertainties of
669 the current calibration conditions (3.4) and to the dependence of the reference meter on these
670 conditions.

671 The following uncertainties shall be evaluated. The evaluation can be made on the basis of
672 measurements or estimations, or a mixture of both. The calculation of uncertainties is
673 described in Annex A. The measurement of dependence on conditions is described in 6.3.1.

- 674 a) Calibration uncertainty of the reference meter. It shall be obtained from its calibration
675 certificate;
- 676 b) Uncertainty due to the change from the conditions in which the reference meter was
677 calibrated and the current calibration conditions u_{change} as calculated in 5.4.3;
- 678 c) Uncertainty due to temperature dependence of the reference meter;
- 679 d) Uncertainty due to dependence on relative humidity of the reference meter. Power meters
680 with integrating sphere are particularly sensitive to absorption peaks of water when using
681 narrow laser sources;
- 682 e) Uncertainty due to dependence on the beam geometry of the reference meter;
- 683 f) Uncertainty due to dependence on multiple reflections. Multiple reflections may exist
684 between the optical input port and the radiation source (for example a connector-adapter
685 combination). Different artefacts will change the measured power;
- 686 g) Uncertainty due to wavelength dependence of the reference meter;
- 687 h) Uncertainty due to dependence on source spectral bandwidth of the reference meter;
- 688 i) Uncertainty due to dependence on state of polarization of the reference meter, except if
689 unpolarized or depolarized light is used for calibration;
- 690 j) Uncertainty due to optical interference. Fabry-Perot cavities can occur between the surface
691 of the detector, of the window and the end of the connector, if used;
- 692 k) Uncertainty due to the resolution of the reference meter. If the resolution of the reference
693 meter is δy_{ref} , the standard uncertainty is (see GUM, F.2.2.1):

$$694 \quad u_{\text{ref,resolution}} = \frac{1}{2\sqrt{3}} \delta y_{\text{ref}} \quad (19)$$

695 l) Uncertainties due to other dependences of the reference meter. Depending on the type of
 696 reference meter, there may be other uncertainties of the reference meter. These should
 697 also be measured or estimated;

698 m) Uncertainty due to ageing of the reference meter.

699 Then calculate the combined standard uncertainty of the reference meter from the n above
 700 standard uncertainties:

$$701 \quad u_{\text{ref}} = \sqrt{\sum_{i=1}^n u_{\text{ref},i}^2 + u_{\text{change}}^2} \quad (20)$$


702 where u_{change} is the uncertainty due to the change of conditions, as determined from 5.4.3.

703 5.4.3 Correction factors and uncertainty caused by the change of conditions

704 The reference meter may exhibit a different response because it was calibrated under
 705 conditions different from the current calibration conditions. Examples for differences between
 706 the two sets of measurement conditions are: Parallel beam versus divergent beam, different
 707 source spectra, a non-reflecting setup versus a setup with multiple reflections, or a large time
 708 span between the two reference dates resulting in ageing of the standard.

709 If the conditions under which the reference meter was calibrated are nominally identical to the
 710 current calibration conditions (their uncertainties can be different) and if the ageing of the
 711 reference meter is negligible, this clause can be skipped ($CF_{\text{change}} = 1$).

712 As indicated in Figure 4, each change comprises the nominal change of condition and the
 713 change of uncertainty.



714

IEC 1829/05

715 **Figure 4 – Change of conditions and uncertainty**

716 For each of the potential error contributions (5.4.3.1 to 5.4.3.7) one should decide if it is
 717 sensible to calculate a correction factor or not. **Alternative A** includes the calculation of a
 718 correction factor with the result of a relatively small uncertainty. **Alternative B** means waiving
 719 the correction factors (or $CF_{\text{change}} = 1$) and taking larger uncertainties into account to embrace
 720 the worst-case conditions.

721 If the alternative A is chosen, the (cumulative) correction factor is:

$$722 \quad CF_{\text{change}} = \frac{r_{\text{previous}}}{r_{\text{current}}} \quad (21)$$

$$723 \quad \text{or} \quad CF_{\text{change}} = 1 - \Delta r \quad (22)$$

724 where

725 r_{previous} is the response of the reference with excitation at the conditions at which it was
726 calibrated;

727 r_{current} is the response of the reference with excitation at the current calibration conditions;

728 Δr is the relative change of response $\Delta r = (r_{\text{current}} - r_{\text{previous}}) / r_{\text{current}}$.

729 Calculate the (cumulative) reference meter's change-related correction factor by accumulating
730 the partial correction factors, $CF_{\text{change},i}$, outlined in 5.4.3.1 to 5.4.3.7. For each influencing
731 quantity X_i , start with the calculation of the partial correction factor:

$$732 \quad CF_{\text{change},i} = 1 - \Delta r_i \quad (23)$$

733 The relative change of response Δr_i can be directly measured by changing the influencing
734 quantity from the “previous” to the “current” calibration conditions or calculated from the
735 nominal change of the influencing quantity Δx_i , and the reference meter's nominal relative
736 dependence on this quantity:

$$737 \quad CF_{\text{change},i} = 1 - c_i \times \Delta x_i \quad (24)$$

738 where c_i is the partial derivate of the relative response on the influence quantity X_i , called
739 **sensitivity coefficient**. See GUM (5.1.3 and 5.1.4) for more detail.

740 If the sensitivity coefficient is not known very well, the following type B uncertainty should be
741 taken into account:

$$742 \quad u_{\text{change},i} = u(c_i) \times \Delta x_i \quad (25)$$

743 where $u(c_i)$ is the standard uncertainty of the sensitivity coefficient. The measurement of the
744 dependences is discussed in 6.3.

745 Finally, calculate the reference meter's cumulative correction factor from the above
746 contributions:

$$747 \quad CF_{\text{change}} = \prod_{i=1}^n CF_{\text{change},i} \quad (26)$$

748 and the combined standard uncertainty due to the change of calibration conditions:

$$749 \quad u_{\text{change}} = \sqrt{\sum_{i=1}^n u_{\text{change},i}^2} \quad (27)$$

750 This correction factor corresponds to a **known** change of response of the reference meter
751 caused by the two different sets of measurement conditions. It is a correction factor to apply to
752 the power read by the reference meter (see 5.3).

753 **5.4.3.1 Correction factor due to temperature change**

754 The correction factor $CF_{\text{change},\Theta}$ should be calculated with the help of the nominal change
755 between the "previous" and the "current" temperature $\Delta\Theta$ and the temperature sensitivity
756 coefficient c_{Θ} of the reference meter (for example in %/°C).

$$757 \quad CF_{\text{change},\Theta} = 1 - c_{\Theta} \times \Delta\Theta \quad (28)$$

758 **5.4.3.2 Correction factor due to change of power level** **C**

760 The uncertainty should be calculated from the nonlinearity of the reference meter between the
761 "previous" and the "current" power level. If necessary, a correction factor can be calculated
762 from:

$$763 \quad CF_{\text{change,NL}} = 10^{\frac{-NL}{10}} \quad (29)$$

764 where NL is the nonlinearity, expressed in decibels (dB). Measurement of nonlinearity is
765 described in Clause 7.

766 **5.4.3.3 Correction factor due to change of beam geometry**

767 The correction factor should be calculated from the change of response measured when
768 changing the beam geometry.

769 **5.4.3.4 Correction factor due to dependence on multiple reflections**

771 The reference meter's optical input port should generally be assumed to be reflective. Such a
772 reflection will travel back to the radiation source, be reflected again, and finally increase the
773 displayed optical power level. This effect will give rise to a correction factor (usually < 1) and
774 an increased uncertainty.

775 If, for example, the source used in the calibration of the reference meter was non-reflective and
776 the source used in the calibration of the test meter is reflective (caused by an optical
777 connector), then the total power indicated by the reference meter is erroneous by the
778 secondary reflection. If one assumes that the secondary reflection contributes an additional
779 5 % of the total power, then the individual correction factor is 0,95. This type of error can be
780 reduced by using sources with highly absorptive enclosures, respectively sources with low-
781 reflectivity connector-adapter combinations.

782 **5.4.3.5 Correction factor due to wavelength change**

783 The correction factor should be calculated with the help of the nominal change of wavelength
784 $\Delta\lambda$ and the reference meter's nominal wavelength dependence c_{λ} .

$$785 \quad CF_{\text{change},\lambda} = 1 - c_{\lambda} \times \Delta\lambda \quad (30)$$

786 **5.4.3.6 Correction factor due to spectral bandwidth change**

787 The correction factor should be calculated with the help of the nominal change of spectral
788 bandwidth and the reference meter's nominal dependence on the spectral bandwidth. Note that
789 the correction factor remains 1 as long as the (uncorrected) wavelength-dependence is **linear**
790 within the spectral bandwidth of the source. In the case that the wavelength dependence is
791 curved, the correction factor can be computed with the help of the wavelength-dependence of
792 the reference meter and the spectra of the two sources used in the calibration of the reference
793 meter and in the calibration of the test meter.

794 **5.4.3.7 Other correction factors**

795 Depending on the type of reference meter and the calibration conditions, there may be other
796 correction factors. These should also be measured or estimated as outlined above.

797 **5.4.4 Uncertainty due to the test meter**

798 Uncertainties arising from the test meter are mainly due to the uncertainties of the calibration
799 conditions and the dependence of the test meter on the conditions. The following uncertainties
800 shall be evaluated. Their determination is similar to 5.4.2. The calculation of uncertainties is
801 described in Annex A, the measurement of dependence on conditions is described in 6.3.1.

- 802 a) Uncertainty due to temperature dependence of the test meter
- 803 b) Uncertainty due to dependence on relative humidity of the test meter. Power meters with
804 integrating sphere are particularly sensitive to absorption peaks of water when using
805 narrow laser sources.
- 806 c) Uncertainty due to dependence on beam geometry. This uncertainty comes from non-
807 uniformity and angle-dependence of the test meter's optical input port.
- 808 d) Uncertainty due to dependence on multiple reflections. Multiple reflections may exist
809 between the optical input port and the radiation source (for example a connector-adapter
810 combination). Different artefacts will change the measured power.
- 811 e) Uncertainty due to wavelength dependence of the test meter
- 812 f) Uncertainty due to dependence on source spectral bandwidth of the test meter
- 813 g) Uncertainty due to dependence on state of polarization of the test meter, except if
814 unpolarized or depolarized light is used for calibration
- 815 h) Uncertainty due to optical interference. Fabry-Perot cavities can occur between the surface
816 of the detector, of the window and the end of the connector, if used
- 817 i) Uncertainty due to the resolution of the test meter. If the resolution of the test meter is
818 δy_{DUT} , the standard uncertainty is (see GUM, F.2.2.1):

$$819 \quad u_{\text{DUT,resolution}} = \frac{1}{2\sqrt{3}} \delta y_{\text{DUT}} \quad (31)$$

- 820 j) Uncertainties due to other dependences of the test meter. Depending on the type of test
821 meter and on the calibration process, there may be other conditions causing uncertainties.

822 Then calculate the combined standard uncertainty contribution of the test meter from the n
823 above standard uncertainties:

$$824 \quad u_{\text{DUT}} = \sqrt{\sum_{i=1}^n u_{\text{DUT},i}^2} \quad (32)$$

825 5.5 Reporting the results

826 The results of each calibration should be reported as required by ISO/IEC 17025. Calibration
827 certificates or calibration reports referring to this standard shall at least include the following
828 information:

- 829 a) all calibration conditions (3.4) as described in 5.2;
- 830 b) the test meter's correction factor(s) (3.6) or deviation(s) (3.8), if the test meter was not
831 adjusted;
- 832 c) on receipt correction factors or deviations and after adjustment (3.2) correction factors or
833 deviations in the case that an adjustment was carried out;
- 834 d) the calibration uncertainty in the form of an expanded uncertainty as described in 5.4;
- 835 e) the instrument state (3.10) of the test meter during the calibration;
- 836 f) evidence that the measurements are traceable (see ISO/IEC 17025:2005, 5.10.4.1 c)).

837 6 Measurement uncertainty of a calibrated power meter

838 6.1 Overview

839 The measurement uncertainty of a calibrated power meter is larger than its calibration
840 uncertainty. It is the combination of the calibration uncertainty and of uncertainty contributions
841 due to the dependence of the power meter on the conditions of measurement.

842 The determination of the measurement uncertainty of a calibrated power meter used at
843 reference conditions or at operating conditions is not part of the calibration process. It is
844 performed for example by manufacturers of power meters in order to establish specifications. It
845 is not mandatory for calibration certificates or calibration reports referring to this standard.

846 6.2 Uncertainty at reference conditions

847 *Reference conditions* (3.25) are used for testing the performance of a power meter or for
848 intercomparisons. They are usually defined by manufacturers in order to specify the smallest
849 uncertainty of a measuring instrument; therefore, they are often identical or close to its
850 calibration conditions.

851 The uncertainty at reference conditions is the uncertainty on the result of a measurement taken
852 by the calibrated and adjusted power meter when operated at reference conditions. It depends
853 on the calibration uncertainty of the power meter, the reference conditions and the dependence
854 of the power meter on the reference conditions. This is the reason why the uncertainty at
855 reference conditions is always larger than the calibration uncertainty. Even when the reference
856 conditions are identical with the calibration conditions (no uncertainty due to change of
857 conditions), the test (power) meter's dependences on the reference conditions have to be
858 added (in quadrature) to the calibration uncertainty for a second time. Calculating the
859 uncertainty at reference conditions of the calibrated test meter is similar to calculating the
860 measurement uncertainty at calibration conditions of the reference meter described in 5.4.2:

$$861 \quad u_{\text{DUT,ref_conditions}} = \sqrt{u^2(CF_{\text{DUT}}) + u_{\text{DUT}}^2} \quad (33)$$

862 where

863 $u(CF_{\text{DUT}})$ is the calibration uncertainty of the test meter, as determined from 5.4, and

864 u_{DUT} is the uncertainty due to the dependence of the test meter on the reference
865 conditions, as determined from 5.4.4.

866 The description of the reference conditions should be made in the same way as the calibration
867 conditions described in 5.2.

868 6.3 Uncertainty at operating conditions

869 The uncertainty at operating conditions (or operating uncertainty, see 2.2.11 of IEC 60359) is
 870 the uncertainty on the result of a measurement taken by the calibrated and adjusted power
 871 meter when operated within a range of operating conditions (3.18). It depends on the
 872 calibration uncertainty, the operating conditions and the dependence of the power meter on the
 873 operating conditions:

$$874 \quad u_{\text{DUT,operating}} = \sqrt{u^2(CF_{\text{DUT}}) + u_{\text{extension}}^2} \quad (34)$$

875 where

876 $u(CF_{\text{DUT}})$ is the calibration uncertainty of the test meter, as determined from 5.4, and
 877 $u_{\text{extension}}$ is the extension uncertainty, due to the dependence of the meter on the operating
 878 conditions, as determined from Equation (38).

879 Contrary to the *calibration conditions* (3.4) described in 5.2, each operating condition should be
 880 described by a range when possible. The set of operating conditions are specified by:

- 881 a) the maximum time span between recalibrations;
- 882 b) the range of ambient temperatures;
- 883 c) the range of power levels (measuring range);
- 884 d) the range of beam geometries described by their spot diameter and numerical aperture, or
 885 the range of fibre types;
- 886 e) the applicable connector-adapter combinations, if any;
- 887 f) the range of wavelengths of the source;
- 888 g) the maximum spectral bandwidth of the source.

889 All possible polarization states are included in the operating conditions by default. A relative
 890 humidity below the condensation point is also assumed.

891 The above conditions may be defined either by the power meter manufacturer or by the
 892 calibration laboratory in charge of the calibration for operating conditions.

893 To calculate the extension uncertainty, combine all uncertainties due to the dependences on
 894 the conditions:


$$895 \quad u_{\text{extension}} = \sqrt{\sum_{i=1}^n u_{\text{extension},i}^2} \quad (35)$$

896 where:

897 $u_{\text{extension},i}$ are contributions to the extension uncertainty;
 898 n is the total number of contributions.

899 6.3.1 Determination of dependences on conditions

900 Each dependence should be recorded as relative change of the meter's response, caused by
 901 changing the relevant condition within its *operating range*. During the test, all other conditions
 902 should be kept at the *calibration conditions*. The zero point is defined by the response at
 903 calibration conditions. This way, each dependence can be specified by a range that is defined
 904 by the maximum positive and negative changes of the response. An asymmetric range about
 905 the zero-point is the usual result as shown in Figure 5.



906

907

Figure 5 – Determining and recording an extension uncertainty

908 In order to obtain good measurement accuracy, the guidelines in Clause 4 should be observed.
 909 Uncertainties in the measurements should be as small as possible, because the measurement
 910 results shall include these uncertainties. It is acceptable to use estimations, instead of
 911 measurements, if these estimations are based on known physical relations or on a sufficiently
 912 large number of characterizing measurements of the same type of test meter.

913 For the determination of the combined standard uncertainty of the test meter at operating
 914 conditions, the limits quantifying the individual dependences shall be converted to standard
 915 uncertainties using Equation (A.5).

916 The individual uncertainties are usually assumed to be independent. However, in some
 917 instances an uncertainty may be strongly dependent on more than one condition. Examples are
 918 outlined in 6.3.4, 6.3.6 and 6.3.7. If the extension uncertainty is substantially increased by
 919 changing the other conditions (within their specified *operating ranges*), this larger uncertainty
 920 shall be recorded. The calculation of the uncertainty shall then be based on these larger
 921 uncertainties.

922 6.3.2 Ageing

923 Ageing is the relative change of response during a period. It can be determined from the
 924 results of successive calibrations of the meter at the same conditions or from indications of the
 925 manufacturer.

926 For a manufacturer, the relative change of response during a period shall be determined with
 927 the assumption of careful use of the instrument. It is recommended to expose the power meter
 928 to its typical environmental conditions, for example ambient temperature $(23 \pm 1) ^\circ\text{C}$ for a
 929 laboratory-type instrument, optical input port non-irradiated, continuously repetitive cycles of
 930 power-on 12 hours, power-off 12 hours, with a total test time equal to the period. The change
 931 of response should be measured by comparison with a working standard. Regular and
 932 traceable recalibration of the working standard will be necessary, in order to exclude ageing of
 933 the working standard. As always, the measurement uncertainty, in this case mostly the
 934 uncertainty of the working standard, shall be taken into account.

935 It is recommended to calculate the ageing uncertainty from a rectangular distribution obtained
 936 as described above (see Clause A.3). If, for example, a detector is known to increase its
 937 response by a maximum of 0,1 % per year at a certain wavelength, then the ageing uncertainty
 938 is characterized by a rectangle that extends from 0 % (at time 0) to +0,1 % (at time 1 year).

939 6.3.3 Dependence on temperature

940 The relative change of response against the response at the calibration conditions should be
 941 measured by changing the temperature within the operating temperature range. The
 942 rectangular uncertainty distribution is then defined by the most negative and the most positive

943 relative changes of the response. Only the extremes of the response as function of the
944 temperature are relevant, not the responses at the extremes of the temperature (see Figure 5).

945 Note that the temperature dependence of the spectral responsivity of semiconductor detectors
946 depends on the wavelength.

947 **6.3.4 Dependence on the power level (nonlinearity)**

948 The relative change of response against the response at the calibration power level should be
949 measured following Clause 7.


950 **6.3.5 Dependence on the type of fibre or on the beam geometry**

951 Fibre-optic power meters may be designed to accept fibres or open beams. It is assumed that
952 the response of the power meter depends on the geometry of the light beam because, for
953 example, of non-uniformity and angle-dependence of the meter's optical input port.

954 The relative change of response should be **measured** with a working standard that exhibits:

- 955 – negligible angle-dependence,
- 956 – negligible surface reflections, and
- 957 – a sufficiently large active area to capture the fibre beams or the open beams.

958



IEC 1831/05

959

960 **Figure 6 – Possible subdivision of the optical reference plane into 10 x 10 squares,**
961 **for the measurement of the spatial response**

962 Another possibility is evaluating the uncertainties with a **mathematical** analysis, based on the
963 assumption that all uncertainties are caused by non-uniform spatial responses of the test
964 meter's reference plane. In preparation of this, the active area of the optical reference plane
965 should be subdivided into an array of squares, for example, 10 x 10 squares as in Figure 6.

966 Then two types of measurements should be carried out:

- 967 a) measurements of the spatial power density, together with the angles of incidence, on the
968 optical reference plane as generated by the applicable beam geometries;
- 969 b) measurements of the test meter's spatial response, weighted with appropriate multipliers
970 which characterize the meter's dependence on oblique incidence (angle dependence), on
971 the test meter's reference plane. The spatial response should be measured with a beam
972 diameter equal to the length of the square.

973 The change of response upon changing the beam parameters can then be evaluated on the
974 basis of **modelling** the necessary measurement results, by multiplying the (spatial) power
975 levels with the spatial responses and adding all products. Note that the spatial responses are
976 usually wavelength-dependent.

977 **6.3.5.1 Measurement of the fibre dependence**

978 In the test of fibre-related uncertainties, multimode fibres under test should be slightly
979 underfilled (see 5.2). The fibres should be terminated by the connector-adapter combination
980 defined by the calibration conditions. Both the connector and the adapter should exhibit a low
981 reflectivity, so that multiple reflections between the connector-adapter combination and the
982 detector do not influence the measurement results. The spectral bandwidth of the source
983 should be narrow enough to avoid averaging over a wide range of wavelengths.

984 Step 1: the output of the reference fibre is measured with both the working standard and the
985 test meter, and the difference is (mathematically) adjusted to zero.

986 Step 2: the above procedure is applied to:

- 987 a) a standard single-mode fibre as defined by IEC 60793-2, and
- 988 b) the (specified) fibre with the largest core diameter, the fibre with the largest
989 numerical aperture or both.

990 The intention of the test is to measure the dependence of the test meter on the type of fibre
991 and on the mode volume. The largest relative change of response against step 1 (positive and
992 negative) should be used to determine the fibre-related uncertainty. The uncertainty shall also
993 include the uncertainty in measuring the fibre outputs with the working standard, caused for
994 example by the effects of non-uniformity, beam divergence and multiple reflections on the
995 working standard.

996 In these measurements, a significant type A uncertainty may be caused by "speckles", in
997 conjunction with the non-uniformity of the optical input port. Speckles are irregular irradiance
998 distributions caused by interference between different modes in a multimode fibre. This effect
999 occurs particularly when the fibre is excited by the (highly coherent) radiation from a laser
1000 diode. This uncertainty can be reduced by averaging a series of measurement results, in which
1001 each sample is taken after a slight movement of the fibre. Fibre movement will change the
1002 speckle pattern. Note this may be accompanied by a change of the total radiant power,
1003 because of a change of the reflected power and the laser diode sensitivity to reflected power.

1004 Speckles do not exist in single-mode fibres when the exciting wavelength is sufficiently longer
1005 than the fibre's cut-off wavelength. Another possibility of eliminating the speckle pattern is
1006 using a less coherent source, such as a filtered LED or a filtered "white" radiation source.

1007 **6.3.5.2 Measurement of open-beam dependence**

1008 Similar to measuring the fibre dependence, the dependence on the spot diameter (3.31) and
1009 the numerical aperture (3.17) of an open beam can be evaluated by comparison with a working
1010 standard that exhibits a uniform large area detector and negligible angle dependence.

1011 To address the problem of combined dependence on spot diameter and numerical aperture, it
1012 may be sufficient to evaluate:

- 1013 a) the relative change of response (against the response at calibration conditions) due to
1014 excitation with the specified smallest spot diameter – smallest numerical aperture; and
- 1015 b) the relative change of response due to excitation with the specified largest spot diameter-
1016 largest numerical aperture.

1017 **6.3.6 Dependence on the connector-adapter combination**

1018 This subclause discusses the test meter's dependence on multiple reflections between the
1019 optical input port and the radiation source (for example an optical connector or other
1020 mechanical parts in the beam path between the source and the optical input port). Note that the
1021 reflections may be specular or diffuse.

1022 The relative change of response should be measured with the help of a working standard that
1023 exhibits negligible angle-dependence and surface reflections. The fibre should be the one of
1024 the calibration conditions. It is advisable to hold the fibre end in place during the measurement,
1025 in order to avoid any bending-induced changes of the power level.

1026 Step 1: the reference beam geometry (respectively the reference fibre), together with the
1027 reference connector-adapter combination, is measured with both the working standard
1028 and the test meter, and the difference is (mathematically) adjusted to zero.

1029 Step 2: the above procedure is applied to all specified connector-adapter combinations, by
1030 repeating each connection several times to reduce type A uncertainties. The largest
1031 relative change of response against step 1 (positive and negative) should be used to
1032 determine the uncertainty. The uncertainty shall also include the type B uncertainty in
1033 measuring the various combinations with the working standard, caused for example by
1034 multiple reflections on the working standard.

1035 Referring to the last paragraph of 6.3.1, it also may be necessary to measure the dependence
1036 with the highest-order fibre, as listed in 6.3.5.1. A high-order fibre will create a larger image on
1037 the optical reference plane, and therefore make limitations in the positioning accuracy more
1038 obvious. In this case, an increased dependence should be recorded.

1039 **6.3.7 Dependence on wavelength**

1040 The relative change of spectral response against the response at the calibration wavelength
1041 should be measured. These measurements will normally be carried out using a spectrally
1042 continuous source imaged through a spectrally discriminating instrument, for example a
1043 monochromator or a number of spectral filters. The stray light, that is light not at the selected
1044 wavelength, should be evaluated, in order to ensure accurate measurement results. The
1045 centroidal wavelength(s) and the spectral bandwidth(s) should also be measured. The
1046 bandwidth should be narrow, because a wide bandwidth in conjunction with a strong curvature
1047 of the test meter's wavelength dependence is capable of producing erroneous measurement
1048 results. Note that extremely narrow spectral bandwidth may cause optical interference
1049 problems, that is comb-like wavelength dependence, when the beam path contains one or
1050 more optical resonators.

1051 The beam geometry should be one of the calibration conditions. It may be possible to
1052 substitute a fibre beam using a combination of lenses and apertures. In this case, care should
1053 be taken to match the irradiated spot diameter and position on the optical reference plane with
1054 those achieved using a fibre input. Care should also be taken to ensure that back reflections
1055 from the optical input port do not add uncertainties to the measurement results.

1056 The measurement should be carried out by direct comparison with a working standard by using
1057 the substitution technique. The working standard should have been calibrated for relative
1058 spectral response.


1059 Because of the relatively low power levels in these measurements, *zero adjustment* of both
1060 power meters is essential. If the instrument comprises means of correction, for example a
1061 calibration curve or a table stored in a memory, the relative change of response from the
1062 **corrected** response has to be measured.

1063 Changing the temperature may strongly influence the wavelength-dependence. For example,
1064 the wavelength-dependence of a germanium photodiode at 1 550 nm is much stronger at 0 °C

1065 than at room temperature. In general, the wavelength uncertainty shall be calculated on the
1066 basis of the largest wavelength-dependence, in this case the one at 0 °C.

1067 **6.3.7.1 Dependence on wavelength due to Fabry-Perot type** 1068 **interference**

1069 When using a narrow spectral bandwidth laser ($B \ll 1$ nm), the spectral response can
1070 sometimes vary rapidly with respect with wavelength as depicted in Figure 7. This is usually
1071 caused by Fabry-Perot cavity(ies) in the optical path to the detector. Fabry-Perot cavities can
1072 occur between the two faces of the window in the detector cap, between one face of the
1073 window and the detector itself, or, if a fibre is used, between the end of the fibre and any of the
1074 other surfaces.



IEC 1832/05

1075

1076 **Figure 7 – Wavelength dependence of response due to Fabry-Perot type interference**

1077 In Figure 7, the peak-to-peak variation reaches $\Delta_{dB} = 0,2$ dB ($\Delta_{\%} = 4,6$ %) which is very
1078 important. The standard uncertainty due to optical interference is the standard deviation of the
1079 sine pattern.


$$1080 \quad u_{\text{int}} = \frac{1}{\sqrt{2}} \times \frac{\Delta_{\%}}{2} = 1,6 \quad \% \quad (36)$$

1081 **6.3.8 Dependence on spectral bandwidth**

1082 This dependence increases with the curvature of the detector's wavelength dependence. The
1083 relative change of response as a function of the spectral bandwidth of the source has to be
1084 tested within the specified range of spectral bandwidths. A monochromator can be used to
1085 generate a variable spectral bandwidth; the actual power level should be measured with a
1086 working standard with negligible wavelength-dependence. The spectral-bandwidth dependence
1087 can also be evaluated by mathematical analysis, based on the known spectral response of the
1088 test meter and on the known spectral characteristics of the source.

1089 **6.3.9 Dependence on polarization**

1090 A method of evaluation of the polarization dependent response (3.22) of the test meter is to
1091 measure the response of the meter multiple times at different states of polarization. A stable
1092 light source polarized to nearly 100 % should be used, otherwise use a polarizer after the
1093 source as shown in Figure 8. A polarization controller is used to convert the fixed input
1094 polarization state to all possible output states.



IEC 1833/05

1095

1096

Figure 8 – Measurement setup of polarization dependent response

1097 The source power instability and the loss variation of the polarization controller should be far
 1098 smaller than the polarization dependence of the test meter. This should be verified by replacing
 1099 the test meter with a detector with very low polarisation dependent response.

1100 NOTE The laser sources may react with unstable power when light with varying polarization state is back-
 1101 reflected; therefore, an attenuator or isolator may have to be inserted between the source and the polarization
 1102 controller.

1103 Another PDR measurement method, the matrix method, can be adapted from the polarization
 1104 dependent loss (PDL) measurement method in IEC 61300-3-12, as described in [3].

1105 6.3.10 Other dependences

1106 Depending on the type of test meter, there may be dependences on other parameters. These
 1107 should also be characterized as relative changes of response against the response at the
 1108 calibration conditions.

1109 One example may be including intensity-modulated optical signals into the operating
 1110 conditions, in the form of specifying a range of modulation frequencies and duty cycles, and
 1111 evaluating the type B uncertainty due to the modulation. Be aware that extreme duty cycles are
 1112 capable of saturating the detector, the electronics or both.

1113 7 Nonlinearity calibration

1114 7.1 General

1115 The nonlinearity (3.16) of the power meter should be calibrated to ensure accurate
 1116 measurements at power levels away from the calibration level and for relative measurements
 1117 such as loss and gain measurements. The calibration should be made by increasing and
 1118 decreasing the power level to detect nonlinearities at the boundaries of each amplifier range or,
 1119 whenever possible, to include measurement results at both sides of each range boundaries, in
 1120 order to include nonlinearities at these boundaries. Be aware that the detector nonlinearity is
 1121 dependent on the wavelength. As an example, an InGaAs detector that is linear at 1 310 nm
 1122 and 1 550 nm may be nonlinear at 850 nm.

1123 Several methods are possible. The superposition method is the reference method, as it is the
 1124 most accurate and does not require a reference standard (self-calibrating method). However,
 1125 the used power steps of 2 (about 3 dB) might be too large to detect nonlinearities that might
 1126 appear at amplifier range boundaries. This limitation may be avoided by starting the calibration
 1127 from several reference powers, or by taking separate measurements of the same power level
 1128 on both sides of the amplifier range boundaries.


1129 All methods use sources with selectable power level, for example (stabilized) laser diode
 1130 sources and variable attenuators. The generated power levels should cover the specified
 1131 measuring range (3.13). During the test, the maximum permissible irradiance of the input port
 1132 should be defined by the optical power at the upper end of the measuring range and by single-
 1133 mode fibre excitation.

1134 The power level saturating the detector is dependent on the beam geometry. A small spot
1135 diameter may saturate the detector at lower power than a larger spot diameter.

1136 NOTE Extreme ambient temperatures may increase the nonlinearity. Referring to the statement on "dependence
1137 on more than one operating condition" in 6.3, it may be necessary to additionally measure the nonlinearity at the
1138 extremes of the operating temperature range, and to record an increased uncertainty at operating conditions.

1139 7.2 Nonlinearity calibration based on superposition

1140 Highly accurate nonlinearity calibration is possible with the superposition method (also known
1141 as the addition method) [4-5]. A "fibred" version of the open-beam double aperture method [6-
1142 7] may be used with single-mode fibres. A possible setup is illustrated in Figure 9. The power is
1143 split into two different paths where shutters are located and then recombined on the power
1144 meter under test.



1145

1146

Figure 9 – Nonlinearity calibration based on superposition

1147 Stable, optically-isolated (to reduce sensitivity to reflections) DFB lasers can be used, provided
1148 that the linewidth is broadened to yield an optimized coherence, as with the procedure for the
1149 absolute calibration. The two paths of the setup should have different lengths (around 100 m
1150 for DFB lasers) to avoid (Mach-Zehnder-type) interference fluctuations, and unused branches
1151 of branching devices shall be terminated. The drawback of this method is its higher insertion
1152 loss: typically around 1,5 dB for the first attenuator, 0,5 dB for the first branching device, 1,5 dB
1153 for the second attenuators and about 3,5 dB for the combining branching device for a total of
1154 about 7 dB. For higher power measurements, an optional optical amplifier (like an EDFA for the
1155 1,55 μm band) can be inserted between the source and the first attenuator.

1156 7.2.1 Procedure

- 1157 (1) Set the attenuators in the two paths so that the power measured on the meter is the same
1158 when light is coming from one path or from the other path.
- 1159 (2) Open both shutters and measure the total power from both paths simultaneously: $P_{ab,i}$.
- 1160 (3) Close the shutter on path b and measure the power from the path a: $P_{a,i}$.
- 1161 (4) Close the shutter on path a, open the shutter on path b and measure the power from the
1162 path b: $P_{b,i}$.
- 1163 (5) If the sum of the individual powers is not equal to the total power, there is a nonlinearity:

$$1164 \quad NL_i = 10 \times \log_{10} \frac{P_{ab,i}}{P_{a,i} + P_{b,i}} \quad (\text{dB}) \quad (37)$$

- 1165 (6) Using the first attenuator, attenuate the total power by a factor 2 ($10 \log_{10} 2 \cong 3,01$ dB) to
1166 the level of the individual powers of the preceding step.

- 1167 (7) Repeat steps (2) to (6) through all the desired range.
 1168 (8) At the end, the global nonlinearity is the sum of all the local nonlinearities expressed in
 1169 decibels (dB), starting calculations from the reference power level where the nonlinearity is
 1170 zero (higher order terms are neglected).

$$1171 \quad NL_{\text{global}}(P_n) = -\sum_{i=0}^{n+1} NL_i \quad \text{for } n = -1, -2, -3, \text{ etc.} \quad (38)$$

$$1172 \quad NL_{\text{global}}(P_0) = 0 \quad (\text{reference power})$$

$$1173 \quad NL_{\text{global}}(P_n) = +\sum_{i=1}^n NL_i \quad \text{for } n = 1, 2, 3, \text{ etc.}$$

1174 where

1175 $n < 0$ indicates power levels lower than the reference power;

1176 $n > 0$ indicates power levels higher than the reference power;

1177 NL_i is the local nonlinearity for the i^{th} step ($i = 0$ for the step where P_{ab} is the reference
 1178 power).

1179 The result is a list of global nonlinearities for the whole power range in steps of 3,01 dB:

1180

Table 2 – Nonlinearity

i	$P_{a,i}$ (W)	$P_{b,i}$ (W)	$P_{a,i} + P_{b,i}$ (W)	$P_{ab,i}$ (W)	NL_i (dB)	$NL_{\text{global}}(P_{ab,i})$ (dB)
2					NL_2	$NL_1 + NL_2$
1					NL_1	NL_1
0				P_0	NL_0	0
-1					NL_{-1}	$-NL_0$
-2					NL_{-2}	$-NL_0 - NL_{-1}$

1181

1182 The largest nonlinearity relative to the reference power is:

$$1183 \quad NL_{\text{max}} = \pm \max(|NL_{\text{global}}|) \text{ (dB)} \quad (39)$$

1184 This result of the nonlinearity calibration can be included in the test meter's calibration
 1185 certificate or calibration report described in 5.5. If desired, NL_{max} may be separately reported
 1186 together with its applicable uncertainty as calculated in the next subclause.

1187 7.2.2 Uncertainties

1188 Typical possible uncertainties of this method include all possible power fluctuations during a
 1189 set of the three measurements such as source fluctuations due to drifts or sensitivity to
 1190 changing reflections, instabilities due to interference if the coherence length of the laser is too
 1191 large, polarization sensitivity and resolution of the power meter. These errors for each step are
 1192 cumulative and will add to the errors of the preceding steps.

1193 Another uncertainty is the inequivalence between the individual powers of each step and also
 1194 with the total power of the next step. If the individual powers are not properly balanced, the
 1195 result will not be reliable. For this last reason, the use of the optional attenuator in each path is
 1196 recommended, as showed in Figure 9 (the shutter is usually included in the attenuator). They

1197 allow the power in each path to be balanced at the beginning of the measurements. Another
 1198 version of the setup uses this approach, but employs two separate laser sources directly
 1199 connected to the second and third attenuators respectively. It has the advantage to start
 1200 measurements at higher powers but it requires communication with the test meter to adjust the
 1201 attenuators at each step.

1202 Calculate first the combined standard uncertainty for the local nonlinearity (one step) $u(NL_i)$ by
 1203 root-sum-squaring all relevant standard uncertainty contribution. Then calculate the standard
 1204 uncertainty of the global nonlinearity with:

$$1205 \quad u(NL_{\text{global}}) = \sqrt{n} \times u(NL_i) \text{ (dB)} \quad (40)$$


1206 where n is the number of 3,01 dB steps counted from the reference level.

1207 7.3 Nonlinearity calibration based on comparison with a calibrated power meter

1208 One possible measurement method is direct comparison of the test meter with a reference
 1209 meter by using the substitution technique. The reference meter is used to determine the output
 1210 power. Then the reference meter is replaced by the test meter. The measurement results of
 1211 both meters are recorded. In this case, errors can be due to the repeatability of the attenuator,
 1212 its PDL, the source power stability and the nonlinearity of the reference meter. The nonlinearity
 1213 of the reference meter should have been calibrated using a more accurate method.

1214 It is advisable to repeat the measurements with the working standard to check for drifts in the
 1215 measurement. In order to extend the measurements to low power levels it is recommended that
 1216 the reference meter incorporates a low noise detector.

1217 Instead of the substitution, simultaneous excitation of both the standard and the test meter,
 1218 with the help of an appropriate beam splitter or branching device is also possible. A beam
 1219 splitter/branching device with an asymmetric ratio, or the use of a second attenuator, will allow
 1220 an extension of the dynamic range of the measurement into both directions. The dependence
 1221 of the ratio to the power level and polarization has to be investigated.



IEC 1835/05

1222

1223 **Figure 10 – Measurement setup for nonlinearity calibration by comparison**

1224 7.3.1 Procedure

- 1225 (1) Set the desired reference power with the help of the first attenuator.
- 1226 (2) Measure the *radiant power* with the reference meter $P_{\text{ref},0}$ and with the test meter $P_{\text{DUT},0}$.
- 1227 (3) Increase (or decrease) the power with the help of the first attenuator and record the power
 1228 measured with the reference meter $P_{\text{ref},i}$ and with the test meter $P_{\text{DUT},i}$.

1229 (4) Calculate the nonlinearity:

$$1230 \quad NL_i = 10 \times \log_{10} \frac{P_{\text{DUT},i}}{P_{\text{DUT},0}} - 10 \times \log_{10} \frac{P_{\text{ref},i}}{P_{\text{ref},0}} \quad (\text{dB}) \quad (41)$$

1231 (5) Repeat steps (3) to (4) to cover the measurement range.

1232 7.3.2 Uncertainties

1233 Possible sources of measurement uncertainties are given in the following list, it may not be
1234 complete. Additional contributions may have to be taken into account, depending on the
1235 measurement setup and procedure. The mathematical basis, Annex A, should be used to
1236 calculate and state the uncertainties.

- 1237 a) nonlinearity of the linearity standard (usually calibrated by the superposition method);
- 1238 b) source instability (back-reflections may cause source instability);
- 1239 c) optical interference (the coherence length of the source should be smaller than the
1240 distance between reflection points);
- 1241 d) polarization dependence of the components;
- 1242 e) resolution of the test meter;
- 1243 f) stability of the ratio of the beam splitter or the branching device if used;
- 1244 g) depending on the procedure, the repeatability of the attenuator.

1245 7.4 Nonlinearity calibration based on comparison with an attenuator

1246 The simplest but least accurate method to measure the nonlinearity, is based on varying the
1247 power level with a calibrated attenuator. The traceability chain (3.34) of the attenuator shall be
1248 determined. Care shall be taken in the calculation of the uncertainty since the calibration of the
1249 attenuator is itself based on the linearity of a calibrated power meter. This method does not
1250 require a second power meter; instead, the reference power levels can then be calculated with
1251 the known attenuation of the attenuator. The main errors arise from the nonlinearity of the
1252 variable attenuator, its PDL in the case of single-mode fibres and the source power stability. Be
1253 also aware of the attenuator's repeatability and wavelength-dependence. This method is
1254 nevertheless useful when high accuracy is not necessary because it is simple and because the
1255 low insertion loss (only the loss of the attenuator) permits to measure at higher power than
1256 other methods (up to the maximum input power at which the attenuator remains linear).

1257 7.5 Calibration of power meter for high power measurement

1258 Most photoelectric detectors become nonlinear above an optical power of about 10 mW.
1259 Sensors designed to measure power at higher power usually incorporate an attenuator in front
1260 of the detector.

1261 Absolute power calibration at high power [8] is not available widely. When not possible, it is
1262 then necessary to calibrate the nonlinearity of the power meter up to high power. In this
1263 context, high power is defined as powers greater than 10 mW. It is not straightforward to use
1264 the same setup as described in the previous clauses since several components may exhibit
1265 nonlinear effects. The behaviour at high power of all elements in the calibration setup
1266 (connectors, attenuators, branching devices, etc.) should be investigated. The superposition
1267 method is the preferred method, since it does not rely on a reference standard, but notice that
1268 the use of a long length of fibre in one path of the superposition system is not desirable at high
1269 powers due to the possibility of nonlinear effects causing apparent nonlinearity of the power
1270 meter.

1271
1272
1273
1274
1275

Annex A (normative)

Mathematical basis for measurement uncertainty calculations

1276

A.1 General

1277 This annex summaries the form of evaluating, combining and reporting the uncertainty of
1278 measurement. It is based on the "Guide to the expression of uncertainty in measurement"
1279 (GUM). It does not relieve the need to consult this guide for more advice.

1280 This standard distinguishes two types of evaluation of uncertainty of measurement. Type A is
1281 the method of evaluation of uncertainty by the statistical analysis of a series of measurements
1282 on the same measurand. Type B is the method of evaluation of uncertainty based on other
1283 knowledge.

1284

A.2 Type A evaluation of uncertainty

1285 The type A evaluation of standard uncertainty can be applied when several independent
1286 observations have been made for a quantity under the same conditions of measurement.

1287 For a quantity X estimated from n independent repeated observations X_k , the arithmetic mean
1288 is:

$$1289 \quad \bar{X} = \frac{1}{n} \sum_{k=1}^n X_k \quad (\text{A.1})$$

1290 This mean is used as the estimate of the quantity, that is $x = \bar{X}$. The experimental standard
1291 deviation of the observations is given by:

$$1292 \quad s(X) = \left[\frac{1}{n-1} \sum_{k=1}^n (X_k - \bar{X})^2 \right]^{1/2} \quad (\text{A.2})$$

1293 where:

1294 \bar{X} is the arithmetic mean of the observed values;

1295 X_k are the measurement samples of a series of measurements;

1296 n is the number of measurements; it is assumed to be large, for example, $n \geq 10$.

1297 The type A standard uncertainty $u_{\text{typeA}}(x)$ associated with the estimate x is the experimental
1298 standard deviation of the mean:

$$1299 \quad u_{\text{typeA}}(x) = s(\bar{X}) = \frac{s(X)}{\sqrt{n}} \quad (\text{A.3})$$

1300 **A.3 Type B evaluation of uncertainty**

1301 The type B evaluation of standard uncertainty is the method of evaluating the uncertainty by
1302 means other than the statistical analysis of a series of observations. It is evaluated by scientific
1303 judgement based on all available information on the variability of the quantity.

1304 If the estimate x of a quantity X is taken from a manufacturer's specification, calibration
1305 certificate, handbook, or other source and its quoted uncertainty $U(x)$ is stated to be a multiple
1306 k of a standard deviation, the standard uncertainty $u(x)$ is simply the quoted value divided by
1307 the multiplier.

$$1308 \quad u(x) = U(x) / k \quad (\text{A.4})$$

1309 If only upper and lower limit X_{\max} and X_{\min} can be estimated for the value of the quantity X , a
1310 rectangular probability distribution is assumed.

1311 The standard uncertainty is

$$1312 \quad u(x) = \frac{\left(|X_{\max} - x|, |X_{\min} - x| \right)_{MAX}}{\sqrt{3}} \quad (\text{A.5})$$

1313 The contribution to the standard uncertainty associated with the output estimate y resulting
1314 from the standard uncertainty associated with the input estimate x is

$$1315 \quad u(y) = c \times u(x) \quad (\text{A.6})$$

1316 where c is the **sensitivity coefficient** associated with the input estimate x , that is the partial
1317 derivative of the model function $y(x)$, evaluated at the input estimate x .

$$1318 \quad c = \frac{\partial y}{\partial x} \quad (\text{A.7})$$

1319 The sensitivity coefficient c describes the extent to which the output estimate y is influenced by
1320 variations of the input estimate x . It can be evaluated by Equation (A.7) or by using numerical
1321 methods, that is by calculating the change in the output estimate y due to a change in the input
1322 estimate x from a model function. Sometimes it may be more appropriate to find the change in
1323 the output estimate y due to the change of x from an experiment.

1324 **A.4 Determining the combined standard uncertainty**

1325 The combined standard uncertainty is used to collect a number of individual uncertainties into a
1326 single number. The combined standard uncertainty is based on statistical independence of the
1327 individual uncertainties; it is calculated by root-sum-squaring all standard uncertainties
1328 obtained from type A and type B evaluation:

$$1329 \quad u_c(y) = \sqrt{\sum_{i=1}^n u_i^2(y)} \quad (\text{A.8})$$

1330 where

1331 i is the current number of individual contribution;

1332 $u_i(y)$ are the standard uncertainty contributions;

1333 n is the number of uncertainties.

1334 NOTE It is acceptable to neglect uncertainty contributions to this equation that are smaller than 1/10 of the largest
1335 contribution, because squaring them will reduce their significance to 1/100 of the largest contribution.

1336 When the quantities above are to be used as the basis for further uncertainty computations,
1337 then the combined standard uncertainty, u_c , can be re-inserted into Equation (A.8). Despite its
1338 partially type A origin, u_c should be considered as describing an uncertainty of type B.

1339 **A.5 Reporting**

1340 In calibration reports and technical data sheets, combined standard uncertainties shall be
1341 reported in the form of expanded uncertainties, together with the applicable level of confidence.
1342 Correction factors or deviations shall be reported. The expanded uncertainty U is obtained by
1343 multiplying the standard uncertainty $u_c(y)$ by a coverage factor k :

$$1344 \quad U = k \times u_c(y) \quad (\text{A.9})$$

1345 For a level of confidence of approximately 95 %, the default level, then $k = 2$. If a level of
1346 confidence of approximately 99 % is chosen, then $k = 3$. The above values for k are valid under
1347 some conditions, see GUM; if these conditions are not met, larger coverage factors are to be
1348 used to reach these levels of confidence.

1349
1350
1351
1352
1353

Annex B (informative)

Linear to dB scale conversion of uncertainties

1354

B.1 Definition of decibel

1355 The decibel is a submultiple of the bel (1 dB = 0,1 B). This unit is used to express values of
1356 power level on a logarithmic scale. The power level is always relative to a reference power P_0 :

$$1357 \quad L_{P/P_0} = 10 \times \log_{10} \left(\frac{P}{P_0} \right) \text{ (dB)} \quad \text{(B.1)}$$

1358 where P and P_0 are expressed in the same linear units.

1359

B.2 Conversion of relative uncertainties

1360 Similar to the previous definition, relative uncertainties, U_{lin} , or relative deviations, can be
1361 expressed in decibels:

$$1362 \quad U_{\text{dB}} = 10 \times \log_{10} (1 + U_{\%}) \quad \text{(B.2)}$$

1363 Reciprocally, U_{lin} can be expressed in % using:

$$1364 \quad U_{\%} = \left[10^{\left(\frac{U_{\text{dB}}}{10} \right)} - 1 \right] \times 100 \quad \text{(B.3)}$$

1365 For small values of U_{lin} , the first term of the applicable Taylor series can be used. Having:

$$1366 \quad \ln(1+x) = \sum_{n=1}^{\infty} \frac{-1^{n+1}}{n} x^n \quad \text{and} \quad \log_{10} x = \frac{\ln x}{\ln(10)} \quad \text{(B.4)}$$

1367 that leads to:

$$1368 \quad U_{\text{dB}} = \frac{10}{\ln(10)} \sum_{n=1}^{\infty} \frac{-1^{n+1}}{n} U_{\text{lin}}^n \approx \frac{10}{\ln(10)} U_{\text{lin}} \quad \text{(B.5)}$$

1369 and two useful expressions:

$$1370 \quad U_{\text{dB}} \approx 4,34 \times U_{\text{lin}} \Leftrightarrow U_{\text{lin}} \approx 0,23 \times U_{\text{dB}} \quad \text{(B.6)}$$

1371

1372

Bibliography

- 1373 [1] IEC 61040:1990, *Power and energy measuring detectors, instruments and equipment for*
1374 *laser radiation*
- 1375 [2] VAYSHENKER, I., LI, X., KEENAN, DA., and SCOTT, TR. Errors due to connectors in
1376 optical fiber power meters. *Technical Digest, Symposium on Optical Fiber Measurement,*
1377 *1996, pp. 49-52.*
- 1378 [3] IVES, DJ. Measurement of Polarisation Dependent Responsivity of optical fibre power
1379 meters and optical spectrum analysers. *Technical digest of the 6th Optical Fibre*
1380 *Measurement Conference, pp. 13-16, 2001.*
- 1381 [4] BUDDE W. Multidecade linearity measurements on Si photodiodes. *Applied Optics, 15*
1382 *May 1979, Vol. 18, No. 10, pp. 1555-1558.*
- 1383 [5] BOIVIN, LP. Automated Absolute and Relative Spectral Linearity Measurements on
1384 Photovoltaic detectors. *Metrologia, 1993, 30, pp. 355-360.*
- 1385 [6] VAYSHENKER, I., YANG, S., LI, X., SCOTT, TR. and CROMER, CL. *Optical Fiber Power*
1386 *Meter Nonlinearity Calibrations at NIST. NIST Special Publication, August 2000, 250-56.*
- 1387 [7] CORREDERA, P., HERNÁNZ, ML., CAMPOS, J., FONTECHA, JL., PONS, A. and
1388 CORRÓNS, A. Application of an addition method to obtain the non-linearity of optical
1389 fibre instrumentation. *Technical Digest of the 4th Optical Fibre Measurement Conference,*
1390 *1997, pp. 146-149.*
- 1391 [8] HALL, SRG., JONES, TCE., and RODDIE, AG. Traceability for high power fibre optic
1392 measurements. *Technical Digest of the 6th Optical Fibre Measurement Conference,*
1393 *2001, pp. 77-80.*
- 1394 *Unnumbered references:*
- 1395 IEC 60050-731:1991, *International Electrotechnical Vocabulary (IEV) – Chapter 731: Optical*
1396 *fibre communication*
- 1397 IEC 60793-1-1, *Optical fibres – Part 1-1: Measurement methods and test procedures – General*
1398 *and guidance*
- 1399 IEC 60793-1-43, *Optical fibres – Part 1-43: Measurement methods and test procedures –*
1400 *Numerical aperture measurement*
- 1401 IEC 60825-1, *Safety of laser products – Part 1: Equipment classification, requirements and*
1402 *user's guide*
- 1403 IEC 60825-2, *Safety of laser products – Part 2: Safety of optical fibre communication systems*
- 1404 IEC 61280-4-1, *Fibre-optic communication subsystem test procedures – Part 4-1: Installed*
1405 *cable plant – Multimode attenuation measurement*
- 1406 IEC TR 61930, *Fibre optic graphical symbology*
- 1407 IEC 61300-3-12, *Fibre optic interconnecting devices and passive components – Basic test and*
1408 *measurement procedures – Part 3-12: Examinations and measurements – Polarization*
1409 *dependence of attenuation of a single-mode fibre optic component: Matrix calculation method*

1410

1411



86/514/CD

COMMITTEE DRAFT (CD)

PROJECT NUMBER: IEC 62129-3 ED1	
DATE OF CIRCULATION: 2017-04-07	CLOSING DATE FOR COMMENTS: 2017-06-30
SUPERSEDES DOCUMENTS: 86/501/CD,86/513/CC	

IEC TC 86 : FIBRE OPTICS	
SECRETARIAT: United States of America	SECRETARY: Mr Steve Swanson
OF INTEREST TO THE FOLLOWING COMMITTEES:	PROPOSED HORIZONTAL STANDARD: <input type="checkbox"/> Other TC/SCs are requested to indicate their interest, if any, in this CD to the secretary.
FUNCTIONS CONCERNED: <input type="checkbox"/> EMC <input type="checkbox"/> ENVIRONMENT <input type="checkbox"/> QUALITY ASSURANCE <input type="checkbox"/> SAFETY	

This document is still under study and subject to change. It should not be used for reference purposes.

Recipients of this document are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

TITLE:
Calibration of wavelength/optical frequency measurement instruments - Part 3:Optical frequency meters internally referenced to a frequency comb

NOTE FROM TC/SC OFFICERS:

CONTENTS

2		
3		
4	FOREWORD	3
5	INTRODUCTION	5
6	1 Scope	6
7	2 Normative references	6
8	3 Terms and definitions	6
9	4 Calibration test requirements	8
10	4.1 Preparation	8
11	4.2 Reference test conditions	8
12	4.3 Traceability	8
13	5 Optical frequency calibration	8
14	5.1 Establishing the calibration conditions	8
15	5.2 Calibration procedure	9
16	5.2.1 General	9
17	5.2.2 Measurement configuration	9
18	5.2.3 Detailed procedure	10
19	5.3 Calibration uncertainty	10
20	5.4 Reporting the results	10
21	Annex A (normative) Mathematical basis	12
22	A.1 General	12
23	A.2 Type A evaluation of uncertainty	12
24	A.3 Type B evaluation of uncertainty	12
25	A.4 Determining the combined standard uncertainty	13
26	A.5 Reporting	14
27	Annex B (informative) Frequency measurement of a stabilized laser with an optical	
28	frequency comb	15
29	Annex C (informative) Frequency-dependence of uncertainty	17
30	Annex D (informative) Examples of stabilized optical frequency comb source	18
31	D.1 Method A (pump pulse source + nonlinear optical effect)	18
32	D.2 Method B (stabilized laser + electro-optical modulator)	18
33	Bibliography	20
34		
35	Figure 1 – Optical frequency meter measurement using a reference source	9
36	Figure 2 – Optical frequency meter measurement using a reference optical frequency	
37	meter	10
38	Figure B.1 – Schematic configuration of optical frequency measurement technique	
39	using an optical comb	15
40	Figure B.2 – Optical spectra of lasers and optical frequency combs	16
41	Figure D.1 – Pump pulse source + nonlinear optical effect	18
42	Figure D.2 – Electro-optical modulator type comb source	19
43		

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**CALIBRATION OF WAVELENGTH /
OPTICAL FREQUENCY MEASUREMENT INSTRUMENTS –****Part 3: Optical frequency meters internally referenced to a frequency
comb**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 62129-3 has been prepared by IEC technical committee 86: Fibre optics.

The text of this standard is based on the technical specification IEC TS 62129-3.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts in the IEC 62129 series, published under the general title *Calibration of wavelength/optical frequency measurements instruments*, can be found on the IEC website.

96 The committee has decided that the contents of this publication will remain unchanged until
97 the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data
98 related to the specific publication. At this date, the publication will be

- 99 • transformed into an International standard,
- 100 • reconfirmed,
- 101 • withdrawn,
- 102 • replaced by a revised edition, or
- 103 • amended.

104

105 A bilingual version of this publication may be issued at a later date.

106

107 The National Committees are requested to note that for this publication the stability date
108 is 2020.

109 THIS TEXT IS INCLUDED FOR THE INFORMATION OF THE NATIONAL COMMITTEES AND WILL BE
110 DELETED AT THE PUBLICATION STAGE.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates
that it contains colours which are considered to be useful for the correct
understanding of its contents. Users should therefore print this document using a
colour printer.

111

112

113

114

115

INTRODUCTION

116 It is essential for realizing fibre optic systems that optical channels are defined in the optical
117 frequency domain, not the wavelength domain. One example: the anchor frequency of the
118 ITU-T grid is 193,1 THz, and the channel spacings of the ITU-T grid are 12,5 GHz, 25 GHz,
119 50 GHz, and 100 GHz [1]¹.

120 ITU-T has also discussed λ -interface systems such as “black link” [2]. “Black link” includes
121 WDM MUX/DEMUX and a transmission fibre, and provides λ -interfaces. Especially in DWDM
122 systems (channel spacing <100 GHz), the uncertainty in specifying optical frequency needs to
123 be minimized.

124 To implement future telecom systems, it is expected that optical frequency measurements will
125 need to be extremely precise. For example, to achieve the channel spacing of 25 GHz, signal
126 optical frequency uncertainty ($U_{f_{\text{sig}}}$) and required measurement uncertainty ($U_{f_{\text{meas}}}$) need to
127 be 2 GHz to 200 MHz ($U_{f_{\text{sig}}}/f = 10^{-5}$ to 10^{-6}) and 200 MHz to 2 MHz ($U_{f_{\text{meas}}}/f = 10^{-6}$ to
128 10^{-8}), respectively. Unfortunately, conventional wavelength meters have measurement
129 uncertainties of 10^{-6} to 10^{-7} . The solution is to use optical frequency measurements since
130 measurement uncertainties can be as small as 10^{-9} , which satisfies the above telecom
131 requirement ($U_{f_{\text{meas}}}/f = 10^{-6}$ to 10^{-8}). Therefore, an optical frequency measurement scheme
132 is necessary for the calibration of future telecom systems.

133 The frequency meter to calibrate with the procedure described in this international standard is
134 the measurement equipment internally utilizing optical frequency comb. In Annex A, the
135 mathematical basis for the uncertainty of measurement is described. The measurement
136 procedure of the frequency with the frequency meter utilizing optical frequency comb is shown
137 in Annex B and the uncertainty depending on the frequency is shown in Annex C. Additionally,
138 the example of the optical frequency comb sources are shown in Annex D.

139 This international standard defines all of the steps involved in the calibration process of the frequency
140 measuring with the optical frequency meter internally utilizing optical frequency comb: establishing the
141 calibration conditions, carrying out the calibration, calculating the uncertainty, and reporting the
142 uncertainty, the calibration conditions and the traceability.

143

¹ Numbers between square brackets refer to the Bibliography.

CALIBRATION OF WAVELENGTH/ OPTICAL FREQUENCY MEASUREMENT INSTRUMENTS –

Part 3: Optical frequency meters internally referenced to a frequency comb

1 Scope

This part of IEC 62129 describes the calibration of optical frequency meters. It is applicable to instruments measuring the optical frequency emitted from sources that are typical for the fibre-optic communications industry. It is assumed that the optical radiation will be coupled to the optical frequency meter by a single-mode optical fibre.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60793-2-50, *Optical fibres – Part 2-50: Product specifications – Sectional specification for class B single-mode fibres*

IEC 60825-1, *Safety of laser products – Part 1: Equipment classification and requirements*

IEC 60825-2, *Safety of laser products – Part 2: Safety of optical fibre communication systems (OFCS)*

IEC TR 61931, *Fibre optic – Terminology*

ISO/IEC Guide 98-3:2008, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

ISO/IEC 17025:2005, *General requirements for the competence of testing and calibration laboratories*

3 Terms and definitions

For the purposes of this document, the terms and definitions contained in IEC TR 61931, as well as the following terms and definitions, apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

3.1 accredited calibration laboratory

calibration laboratory authorized by the appropriate national organization to issue calibration certificates with a minimum specified uncertainty, which demonstrate traceability to national measurement standards

182 **3.2**
183 **calibration**
184 set of operations that establish, under specified conditions, the relationship between the
185 values of quantities indicated by a measuring instrument and the corresponding values
186 realized by measurement standards

187 Note 1 to entry: The result of a calibration permits either the assignment of values of measurands to the
188 indications or the determination of corrections with respect to indications.

189 Note 2 to entry: A calibration may also determine other metrological properties such as the effect of influence
190 quantities.

191 Note 3 to entry: The result of a calibration may be recorded in a document, sometimes called a calibration
192 certificate or a calibration report.

193 [SOURCE: ISO/IEC Guide 99:2007, 2.39, modified – only the first part of the definition is used]
194 [3]

195 **3.3**
196 **national (measurement) standard**
197 measurement standard recognized by a national decision to serve, in a country, as the basis
198 for assigning values to other measurement standards of the quantity concerned

199 [Related to ISO/IEC Guide 99:2007, 5.3]

200 **3.4**
201 **national standards laboratory**
202 laboratory which maintains the national measurement standard

203 **3.5**
204 **reference standard**
205 measurement standard, generally having the highest metrological quality available at a given
206 location or in a given organization, from which measurements made there are derived

207 [Related to ISO/IEC Guide 99:2007, 5.6]

208 **3.6**
209 **traceability**
210 property of the result of a measurement or the value of a measurement standard whereby it
211 can be related to stated references, usually national or international measurement standards,
212 through an unbroken chain of comparisons all having stated uncertainties

213 [Related to ISO/IEC Guide 99:2007, 2.41]

214 **3.7**
215 **traceability chain**
216 unbroken chain of comparison

217 [Related to ISO/IEC Guide 99:2007, 2.42]

218 **3.8**
219 **working standard**
220 measurement standard that is used routinely to calibrate or check measuring instruments

221 Note 1 to entry: A working standard is usually calibrated against a reference standard.

222 [Related to ISO/IEC Guide 99:2007, 5.7].

223 4 Calibration test requirements

224 4.1 Preparation

225 The following recommendations apply.

226 The calibration laboratory should satisfy requirements of ISO/IEC 17025.

227 There should be a documented measurement procedure for each type of calibration
228 performed, giving step-by-step operating instructions and equipment to be used.

229 The environmental conditions shall be commensurate with the degree of uncertainty that is
230 required for calibration:

- 231 a) the environment shall be clean;
- 232 b) temperature monitoring and control is required;
- 233 c) all laser sources shall be safely operated (refer to IEC 60825-1 and IEC 60825-2).

234 Perform all tests at an ambient room temperature of $23\text{ °C} \pm 3\text{ °C}$ with a relative humidity of
235 $(50 \pm 20)\%$ unless otherwise specified. Give the test equipment a minimum of 2 h prior to
236 testing to reach equilibrium with its environment. Allow the optical frequency meter a warm-up
237 period in accordance with the manufacturer's instructions.

238 4.2 Reference test conditions

239 The reference test conditions usually include the following parameters and, if necessary, their
240 tolerance bands: date, temperature, relative humidity, displayed power level, displayed optical
241 frequency, fibre, connector-adapter combination, (spectral) bandwidth and resolution
242 bandwidth (spectral resolution) set. Unless otherwise specified, use a single-mode optical
243 fibre input pigtail as prescribed by IEC 60793-2-50, having a length of at least 2 m.

244 Operate the optical frequency meter in accordance with the manufacturer's specifications and
245 operating procedures. Where practical, select a range of test conditions and parameters
246 which emulate the actual field operating conditions of the optical frequency meter under test.
247 Choose these parameters so as to optimize the optical frequency meter's uncertainties, as
248 specified by the manufacturer's operating procedures.

249 NOTE The calibration results only apply to the set of test conditions used in the calibration process.

250 4.3 Traceability

251 The requirements of ISO/IEC 17025 should be met.

252 Make sure that any test equipment which has a significant influence on the calibration results
253 is calibrated. Upon request, specify this test equipment and its calibration chain(s). The
254 recalibration period(s) shall be defined and documented.

255 5 Optical frequency calibration

256 5.1 Establishing the calibration conditions

257 Establishing and maintaining the calibration conditions is an important part of the calibration,
258 because any change of these conditions is capable of producing erroneous measurement
259 results. The calibration conditions should be a close approximation to the intended operating
260 conditions. This ensures that the (additional) uncertainty in the operating environment is as
261 small as possible. The calibration conditions should be specified in the form of nominal values
262 with uncertainties when applicable. In order to meet the requirements of this standard, the
263 calibration conditions shall at least consist of:

- 264 a) date of calibration;
- 265 b) ambient temperature, with uncertainty, for example $23\text{ °C} \pm 3\text{ °C}$. The temperature may
266 need to be monitored continuously to ensure that it remains within the prescribed limits;
- 267 c) ambient relative humidity, for example 30 % – 70 %. The ambient relative humidity may
268 need to be monitored continuously to ensure that it remains within the prescribed limits. A
269 relative humidity below the condensation point is assumed by default;
- 270 d) input optical power (that has to fall within the allowable specification for the optical
271 frequency meters);
- 272 e) if a transition locked source is used, then the quality of the lock shall be continuously
273 monitored during the measurements; a lock indicator can be sufficient.

274 The above conditions may not be exhaustive. There may be other parameters that have a
275 significant influence on the measurement uncertainty and therefore should be reported, too.

276 5.2 Calibration procedure

277 5.2.1 General


278 The main steps of the calibration procedure are as follows:

- 279 a) establish and record the appropriate measurement conditions (see 5.1 a)~c)). Switch on
280 all instrumentation and wait for enough time to stabilize;
- 281 b) set up the reference source with condition 5.2 d);
- 282 c) set up the instrument state of the test optical frequency meter according to the instruction
283 manual. Select appropriate units;
- 284 d) record the instrument states of the optical frequency meter.

285 5.2.2 Measurement configuration

286 For calibration laboratories, a stabilized laser can be utilized as a reference source.


287 Figure 1 shows the configuration using the reference source. The temperature of the
288 environment is monitored.



290 **Figure 1 – Optical frequency meter measurement using a reference source**

291 Figure 2 shows the configuration using a reference optical frequency meter.

292 It is necessary that the measurements be performed simultaneously on both the reference
293 and the test optical frequency meters.



294

295 1 – Optical frequency meter measurement using a reference optical frequency meter

296 5.2.3 Detailed procedure

297 Typically, 50 samples (n) are taken per measurement unless otherwise defined in the relevant
298 specification.

299 The measurement process is as follows:

- 300 a) allow the equipment to reach equilibrium;
- 301 b) configure the data acquisition software;
- 302 c) ensure that the optical source is locked and is operating correctly;
- 303 d) run the data acquisition software.

304 The correction factor is determined from the difference between the reference optical
305 frequency and the mean values from each measurement:

$$306 \quad CF = f_{ref} - \frac{1}{n} \sum_{i=1}^n f_{test_i} \quad (1)$$

307 where f_{ref} is the reference optical frequency and f_{test} is the optical frequency measured by the
308 test optical frequency meter.

309 5.3 Calibration uncertainty

310 Note that the following list may not be complete. Additional contributions may have to be
311 taken into account, depending on the measurement setup and procedure. The mathematical
312 basis, Annex A, should be used to calculate and state the uncertainty.

- 313 a) Source uncertainty (how well the source is stabilized to the natural standard)
- 314 b) Standard deviation of measurement samples

315 5.4 Reporting the results

316 The results of each calibration should be reported as required by ISO/IEC 17025. Calibration
317 certificates referring to this standard shall at least include the following information:

- 318 a) all calibration conditions of the calibration process as described in 5.1;
- 319 b) the test meter's correction factor(s) or deviation(s), if the test meter was not adjusted;
- 320 c) on receipt, correction factors or deviations and, after adjustment, correction factors or
321 deviations in the case that an adjustment was carried out;

- 322 d) the calibration uncertainty in the form of an expanded uncertainty as described in 5.2 and
- 323 Annex A;
- 324 e) the instrument state of the test meter during the calibration;
- 325 f) evidence that the measurements are traceable (see 5.10.4.1c) of ISO/IEC 17025:2005).

Annex A (normative)

Mathematical basis

A.1 General

This annex summarizes the form of evaluating, combining and reporting the uncertainty of measurement. It is based on ISO/IEC 98-3. It does not replace the need to consult this guide for more advice.

This technical specification distinguishes two types of evaluation of uncertainty of measurement. Type A is the method of evaluation of uncertainty by the statistical analysis of a series of measurements on the same measurand. Type B is the method of evaluation of uncertainty based on other knowledge.

A.2 Type A evaluation of uncertainty

Type A evaluation of standard uncertainty can be applied when several independent observations have been made for a quantity under the same conditions of measurement.

For a quantity X estimated from n independent repeated observations X_k , the arithmetic mean is

$$\bar{X} = \frac{1}{n} \sum_{k=1}^n X_k \quad (\text{A.1})$$

This mean is used as the estimate of the quantity, that is $x = \bar{X}$. The experimental standard deviation of the observations is given by

$$s(X) = \left[\frac{1}{n-1} \sum_{k=1}^n (X_k - \bar{X})^2 \right]^{1/2} \quad (\text{A.2})$$

where

\bar{X} is the arithmetic mean of the observed values;

X_k are the measurement samples of a series of measurements;

n is the number of measurements; it is assumed to be large, for example, $n \geq 10$.

The type A standard uncertainty $u_{\text{typeA}}(x)$ associated with the estimate x is the experimental standard deviation of the mean

$$u_{\text{typeA}}(x) = s(\bar{X}) = \frac{s(X)}{\sqrt{n}} \quad (\text{A.3})$$

A.3 Type B evaluation of uncertainty

Type B evaluation of standard uncertainty is the method of evaluating the uncertainty by means other than the statistical analysis of a series of observations. It is evaluated by scientific judgement based on all available information on the variability of the quantity.

358 If the estimate x of a quantity X is taken from a manufacturer's specification, calibration
 359 certificate, handbook, or other source and its quoted uncertainty $U(x)$ is stated to be a multiple
 360 k of a standard deviation. The standard uncertainty $u(x)$ is simply the quoted value divided by
 361 the multiplier:

$$362 \quad u(x) = \frac{U(x)}{k} \quad (\text{A.4})$$

363 If only upper and lower limit X_{\max} and X_{\min} can be estimated for the value of the quantity X , a
 364 rectangular probability distribution is assumed.

365

366 The standard uncertainty is

$$367 \quad u(x) = \frac{(|X_{\max} - x|, |X_{\min} - x|)_{MAX}}{\sqrt{3}} \quad (\text{A.5})$$

368 The contribution to the standard uncertainty associated with the output estimate y resulting
 369 from the standard uncertainty associated with the input estimate x is

$$370 \quad u(y) = c \times u(x) \quad (\text{A.6})$$

371 where c is the sensitivity coefficient associated with the input estimate x , that is the partial
 372 derivative of the model function $y(x)$, evaluated at the input estimate x .

$$373 \quad c = \frac{\partial y}{\partial x} \quad (\text{A.7})$$

374 The sensitivity coefficient c describes the extent to which the output estimate y is influenced
 375 by variations of the input estimate x . It can be evaluated by Equation (A.7) or by using
 376 numerical methods, i.e. by calculating the change in the output estimate y due to a change in
 377 the input estimate x from a model function. Sometimes it may be more appropriate to find the
 378 change in the output estimate y due to the change of x from an experiment.

379 **A.4 Determining the combined standard uncertainty**

380 Combined standard uncertainty is used to collect a number of individual uncertainties into a
 381 single number. The combined standard uncertainty is based on statistical independence of the
 382 individual uncertainties. It is calculated by root-sum-squaring all standard uncertainties
 383 obtained from type A and type B evaluation.

$$384 \quad u_c(y) = \sqrt{\sum_{i=1}^n u_i^2(y)} \quad (\text{A.8})$$

385 where

386 i is the current number of individual contributions;

387 $u_i(y)$ are the standard uncertainty contributions;

388 n is the number of uncertainties.

389 NOTE It is acceptable to neglect uncertainty contributions to Equation (A.9) that are smaller than 1/10 of the
390 largest contribution because squaring them will reduce their significance to 1/100 of the largest contribution.

391 When the quantities above are to be used as the basis for further uncertainty computations,
392 then the combined standard uncertainty, u_c , can be re-inserted into Equation (A.8). Despite its
393 partially type A origin, u_c should be considered as describing an uncertainty of type B.

394 **A.5 Reporting**

395 In calibration reports and technical data sheets, combined standard uncertainties shall be
396 reported in the form of expanded uncertainties, together with the applicable level of
397 confidence. Correction factors or deviations shall be reported. The expanded uncertainty U is
398 obtained by multiplying the standard uncertainty $u_c(y)$ by a coverage factor k

$$399 \quad U = k \cdot u_c(y) \quad (\text{A.9})$$

400 For a level of confidence of approximately 95 %, the default level, then $k = 2$. If a level of
401 confidence of approximately 99 % is chosen, then $k = 3$. The above values for k are valid
402 under some conditions (see GUM); if these conditions are not met, larger coverage factors are
403 to be used to reach these levels of confidence.

404

Annex B (informative)


Frequency measurement of a stabilized laser with an optical frequency comb

405
406
407
408
409

410 For optical frequency measurement, equally-spaced “frequency comb” lines (spacing of up to
411 50 GHz) from an optical frequency comb are utilized as a “ruler” for optical frequency
412 measurement [4 – 15]. Optical frequency measurements provide more accurate calibration
413 than interferometric wavelength measurements in air by eliminating the effects of refractive
414 indices.

415 Some examples of practical optical frequency comb are shown in Annex B.

416 Figure B.1 is the schematic configuration of an optical frequency measurement technique that
417 uses optical frequency combs.



418

IEC 0579/14

419 **Figure B.1 – Schematic configuration of optical frequency measurement technique**
420 **using an optical comb**

421 Figure B.2 shows the optical spectrum of the laser and the optical frequency comb. The
422 optical frequency comb generates an optical frequency comb with uniform spacing ($f(\text{comb}$
423 spacing)) which is equal to the electrical clock frequency driving the optical frequency comb.
424 $f(\text{comb spacing})$ is also equal to the pulse repetition rate. Thus, the uncertainty of comb
425 spacing is proportional to the uncertainty of the electrical clock frequency. The comb spacing
426 generally lies between 100 MHz and 25 GHz. In this case, the stabilized laser ($f(\text{stabilized}$
427 laser)) output is combined with the optical frequency comb, and then these two lights are
428 input to an optical-electrical (O/E) converter. The beat frequency ($f(\text{beat})$) between the two
429 lights is taken as the output of the O/E converter. The optical frequency ($f(\text{stabilized laser})$) of
430 the stabilized laser can be calculated by Equation (B.1).

$$431 \quad f(\text{stabilized laser}) = f(N) \pm f(\text{beat}) \quad (\text{B.1})$$


432 Here, $f(N)$ is the optical frequency of the N -th mode of optical frequency comb, and is the
433 summation of $f(\text{comb spacing})$ and the carrier envelope offset frequency $f(\text{CEO})$, as shown in
434 Equation (B.2).

$$435 \quad f(N) = N \times f(\text{comb spacing}) + f(\text{CEO}) \quad (\text{B.2})$$

436 Here, N is the large integer, and can be determined with a wavelength meter. The sign of the
 437 beat frequency (+ or –) can be deduced by changing $f(\text{stabilized laser})$ slightly.

438 $f(\text{CEO})$ is related to the pulse-to-pulse phase shift, $\Delta\phi$, between the peak of electrical field
 439 and the peak of envelope [5], as shown in Equation (B.3).

$$440 \quad f(\text{CEO}) = (\Delta\phi / 2\pi) f(\text{comb spacing}) \quad (\text{B.3})$$



441

IEC 0580/14

442 **Key**

443	f	optical frequency
444	$f(\text{stabilized laser})$	stabilized laser frequency
445	$f(\text{comb spacing})$	comb spacing
446	$f(\text{beat})$	beat frequency
447	$f(N)$	optical frequency of the N-th comb mode
448	$f(\text{CEO})$	carrier envelope offset frequency

449

Figure B.2 – Optical spectra of lasers and optical frequency combs

450

Annex C (informative)

Frequency-dependence of uncertainty

451
452
453
454
455 The uncertainty arising from the electrical clocks in Clauses D.1 to D.2 will affect the mode
456 spacing in a way that varies with comb mode number.

457 From Equation (B.1)

$$458 \quad f(\text{tuneable laser}) = f(\text{stabilized laser}) + N \times f(\text{comb spacing}) \pm f(\text{beat})$$

459 Therefore the uncertainty in the comb spacing multiplies up by N , which varies depending on
460 the frequency of the tuneable laser under test.

461 For example, when the uncertainty arising from the electrical clock is less than 10^{-8} and the
462 measurement frequency range is 5 THz, an uncertainty of less than 50 kHz can be realized
463 within the entire measurement frequency range. However, if the uncertainty of the electrical
464 clocks degrades to 10^{-6} , the measurement uncertainty in the frequency of 25 GHz away from
465 the stabilized laser will be 25 kHz; whereas at a 5 THz separation, it will be 5 MHz. In this
466 case, the frequency dependent uncertainty may exceed the measurement uncertainty required
467 for telecom systems (i.e. 200 MHz to 2 MHz) for some frequency values.

468 When the uncertainty of the electrical clocks is not clear, it is recommended to test the device
469 with a minimum of two reference frequencies: one close to the frequency of the stabilized
470 laser, and the second reference at a widely separated frequency. From these results, the
471 frequency dependence of the uncertainty can be calculated.

472

Annex D (informative)

Examples of stabilized optical frequency comb source

D.1 Method A (pump pulse source + nonlinear optical effect)

Figure D.1 shows an optical comb system combining pump pulse source and nonlinear optical effects [4-13, 15]. The pump pulse source is driven by an electrical clock (frequency: $f(\text{clock})$). The pump pulse source generates an optical pulse train with a repetition rate of $f(\text{clock})$. The pump pulse train is amplified when the pulse power is small. The optical pulse train is input to a nonlinear material (fibre, etc.) and the resulting spectral broadening can exceed an octave. By comparing the optical frequencies of the octave-broadened comb with that of the second harmonic, one can stabilize and measure the frequency offset ($f(\text{CEO})$). The frequencies $f(\text{CEO})$ and $f(\text{comb spacing})$ are locked to the electrical clock, thus yielding a very stable optical comb without the use of an external stabilized laser. The uncertainty of the optical comb is determined by that of the electrical clock so very small values can be expected.

The pump pulse source can be realized with several sources. For example, a mode lock laser can be utilized. A titanium-doped sapphire (Ti:S) laser or a fibre laser can be used as the mode-locked laser. These optical frequency comb systems can be extended to any IR wavelength region (1 μm to 2 μm). As second example, the supercontinuum (SC) comb source can be utilized. SC is a spectral broadening phenomenon and is realized when nonlinear materials are pumped by short optical pulses. It occurs due to the combined effects of self-phase modulation, cross-phase modulation, parametric, four-wave mixing and Raman scattering. A superbroadened bandwidth of more than 200 nm (25 THz) with the spacing range from several GHz to several 10 GHz at 1,5 μm and the uncertainty of 1 MHz for 25 GHz spacing in the telecom bands has been reported[15].





Figure D.1 – Pump pulse source + nonlinear optical effect

D.2 Method B (stabilized laser + electro-optical modulator)

Figure D.2 shows an electro-optical modulator type comb source with a stabilized laser. The stabilized laser output is amplified and then input into the electro-optical modulator which is driven by an electrical clock. This method has been reported to offer 5 THz bandwidth optical comb generation with 6,25 GHz spacing [14].



505

IEC 0584/14

506

Figure D.2 – Electro-optical modulator type comb source

507

508

509

510
511

Bibliography

512 International standards:

- 513 [1] ITU-T, G.694.1, *Spectral grids for WDM applications: DWDM frequency grid*
- 514 [2] ITU-T, G.695, *Optical interfaces for coarse wavelength division multiplexing*
515 *applications*
- 516 [3] ISO/IEC Guide 99:2007, *International Vocabulary of metrology – Basic and general*
517 *concepts and associated terms (VIM)*

518 Other references:

- 519 [4] UDEM, Th. et al., *Phys. Rev. Lett.*, 82, 3568-3571 (1999)
- 520 [5] JONES, D.J. et al., *Science*, 288, 635-639 (2000)
- 521 [6] UDEN, T et al., *Nature*, 416, 233-237 (2002)
- 522 [7] http://nobelprize.org/nobel_prizes/physics/laureates/2005/
- 523 [8] HONG, F.-L. et al., *Opt. Lett.*, vol.28, no. 23, pp.2324-2326 (2003)
- 524 [9] EDWARDS, C.S. et al., *Opt. Lett.*, vol.29, no.6, pp. 566-568 (2004)
- 525 [10] SCHIBILI, T. R. et al., *Opt. Lett.*, vol.29, no.21, pp.2467-2469 (2004)
- 526 [11] INABA, H. et al., *Optics Express*, vol.14, no. 12, pp. 5223-5231 (2006)
- 527 [12] NEWBURY, N.R. et al., *J. Opt. Soc. Am. B*, vol. 24, no. 8, pp.1756-1770 (2007)
- 528 [13] PENG, J.-L. et al., *Appl Phys B*, vol.92, pp.513–518 (2008)
- 529 [14] KOUROGI, M. et al., *IEEE J. Quantum Electron.*, 29, 2693 (1993)
- 530 [15] SUZUKI, K. et al., *Electron. Lett.*, vol.40, no,17, pp.1078-1079 (2004)

531 Unnumbered reference:

- 532 IEC 60359, *Electrical and electronic measuring equipment – Expression of performance*



PROJECT NUMBER:
IEC 61753-121-2 ED2

DATE OF CIRCULATION:
2017-03-24

CLOSING DATE FOR VOTING:
2017-05-05

SUPERSEDES DOCUMENTS:
86B/3908/CDV, 86B/3963A/RVC

IEC SC 86B : FIBRE OPTIC INTERCONNECTING DEVICES AND PASSIVE COMPONENTS	
SECRETARIAT: Japan	SECRETARY: Mr Shigeru Tomita
OF INTEREST TO THE FOLLOWING COMMITTEES:	HORIZONTAL STANDARD: <input type="checkbox"/>
FUNCTIONS CONCERNED: <input type="checkbox"/> EMC <input type="checkbox"/> ENVIRONMENT <input type="checkbox"/> QUALITY ASSURANCE <input type="checkbox"/> SAFETY	
<input checked="" type="checkbox"/> SUBMITTED FOR CENELEC PARALLEL VOTING Attention IEC-CENELEC parallel voting The attention of IEC National Committees, members of CENELEC, is drawn to the fact that this Final Draft International Standard (FDIS) is submitted for parallel voting. The CENELEC members are invited to vote through the CENELEC online voting system.	<input type="checkbox"/> NOT SUBMITTED FOR CENELEC PARALLEL VOTING

This document is a draft distributed for approval. It may not be referred to as an International Standard until published as such.

In addition to their evaluation as being acceptable for industrial, technological, commercial and user purposes, Final Draft International Standards may on occasion have to be considered in the light of their potential to become standards to which reference may be made in national regulations.

Recipients of this document are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

TITLE:
Fibre optic interconnecting devices and passive components - Performance standard - Part 121-2: Simplex and duplex cords with single-mode fibre and cylindrical ferrule connectors for category C - Controlled environment

NOTE FROM TC/SC OFFICERS:

CONTENTS

FOREWORD	4
1 Scope	6
2 Normative references	6
3 Terms and definitions	8
4 Description	9
4.1 General.....	9
4.2 Optical fibres	9
4.3 Cable design and construction	9
4.4 Optical connectors	9
4.4.1 Mechanical connectivity.....	9
4.4.2 Optical performance requirements	9
4.4.3 Connector set performance requirements	9
4.5 Cable bend radius.....	9
5 Tests	9
5.1 General.....	9
5.2 Measurement wavelengths.....	10
5.3 Device under test.....	10
5.4 Test report	10
6 Test procedure	10
6.1 General.....	10
6.2 Visual examination.....	10
6.3 Fibre optic connector plug end face	10
6.4 Optical performance requirements	11
6.5 Environmental performance requirements	12
6.6 Mechanical performance requirements.....	13
Annex A (normative) Sample size requirements	15
Annex B (normative) Visual examination of outer cable sheath movement	16
B.1 Scope	16
B.2 Preparation of the DUT and initial visual examination	16
B.3 Final visual examination of outer cable sheath movement.....	16
Annex C (normative) Change of temperature	17
Bibliography.....	18
Figure B.1 – Initial marking of the cable sheath.....	16
Figure B.2 – Final visual examination.....	16
Figure C.1 – Change of temperature test configuration	17
Table 1 – Wavelengths for attenuation and return loss measurements	10
Table 2 – Visual examination requirements.....	10
Table 3 – End face requirements	11
Table 4 – Optical performance requirements.....	11
Table 5 – Environmental performance requirements	12

Table 6 – Mechanical performance requirements 13
Table A.1 – Sample size requirements 15

INTERNATIONAL ELECTROTECHNICAL COMMISSION

FIBRE OPTIC INTERCONNECTING DEVICES AND PASSIVE COMPONENTS – PERFORMANCE STANDARD –

Part 121-2: Simplex and duplex cords with single-mode fibre and cylindrical ferrule connectors for category C – Controlled environment

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61753-121-2 has been prepared by subcommittee 86B: Fibre optic interconnecting devices and passive components, of IEC technical committee 86: Fibre optics.

This second edition cancels and replaces the first edition published in 2010. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) merge an optical performance requirement of a reference cord;
- b) delete Annexes D and E due to updated relevant standard document;

c) modify the whole document structure according to the latest ISO/IEC Directives.

The text of this International Standard is based on the following documents:

FDIS	Report on voting
86B/XX/FDIS	86B/XX/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts of IEC 61753 series, published under the general title *Fibre optic interconnecting devices and passive components – Performance standard*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

The National Committees are requested to note that for this document the stability date is 2025.

THIS TEXT IS INCLUDED FOR THE INFORMATION OF THE NATIONAL COMMITTEES AND WILL BE DELETED AT THE PUBLICATION STAGE.

FIBRE OPTIC INTERCONNECTING DEVICES AND PASSIVE COMPONENTS – PERFORMANCE STANDARD –

Part 121-2: Simplex and duplex cords with single-mode fibre and cylindrical ferrule connectors for category C – Controlled environment

1 Scope

This part of IEC 61753 specifies the test requirements for cords including reference cords used in a controlled (Category C) environment according to IEC 61753-1, where the connectors already comply with the Category C requirements of IEC 61753-1. The tests selected are a subset of the connector tests from IEC 61753-1 appropriate for requalification with additional requirements relevant to cords and the connector/cable interface.

The cords consist of simplex or duplex fibre optic cable terminated at each end of the cable with single-mode fibre optic connector plugs with cylindrical ferrules. The operational wavelength range is between 1 260 nm and 1 625 nm. Short length cords are used as test samples as the attenuation of the cord and the temperature cycling performance will be affected by longer lengths of cable. It is important that any qualification of a cord whose length is greater than 5 m takes these factors into account.

The relevant requirements for the mechanical interface of connector sets are covered by the IEC 61754 all parts. The relevant requirements for the optical interface of connector sets are covered by IEC 61755 (all parts). The relevant requirements for performance of connector sets are covered by IEC 61753 (all parts). The relevant requirements for fibres are covered by IEC 60793-2-50. The relevant requirements for cables for cords are covered by IEC 60794-2-50.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60793-2-50, *Optical fibres – Part 2-50: Product specifications – Sectional specification for class B single-mode fibres*

IEC 60794-2-50, *Optical fibre cables – Part 2-50: Indoor cables – Family specification for simplex and duplex cables for use in terminated cable assemblies*

IEC 60794-2-51, *Optical fibre cables – Part 2-51: Indoor cables – Detail specification for simplex and duplex cables for use in cords for controlled environment*

IEC 61300-1, *Fibre optic interconnecting devices and passive components – Basic test and measurement procedures – Part 1: General and guidance*

IEC 61300-2-4, *Fibre optic interconnecting devices and passive components – Basic test and measurement procedures – Part 2-4: Tests – Fibre/cable retention*

IEC 61300-2-22, *Fibre optic interconnecting devices and passive components – Basic test and measurement procedures – Part 2-22: Tests – Change of temperature*

IEC 61300-2-42, *Fibre optic interconnecting devices and passive components – Basic test and measurement procedures – Part 2-42: Tests – Static side load for strain relief*

IEC 61300-2-44, *Fibre optic interconnecting devices and passive components – Basic test and measurement procedures – Part 2-44: Tests – Flexing of the strain relief of fibre optic devices*

IEC 61300-3-1, *Fibre optic interconnecting devices and passive components – Basic test and measurement procedures – Part 3-1: Examinations and measurements – Visual examination*

IEC 61300-3-3, *Fibre optic interconnecting devices and passive components – Basic test and measurement procedures – Part 3-3: Examinations and measurements – Active monitoring of changes in attenuation and return loss*

IEC 61300-3-6, *Fibre optic interconnecting devices and passive components – Basic test and measurement procedures – Part 3-6: Examinations and measurements – Return loss*

IEC 61300-3-22, *Fibre optic interconnecting devices and passive components – Basic test and measurement procedures – Part 3-22: Examinations and measurements – Ferrule compression force*

IEC 61300-3-25, *Fibre optic interconnecting devices and passive components – Basic test and measurement procedures – Part 3-25: Examinations and measurements – Concentricity of non-angled ferrules and non-angled ferrules with fibre installed*

IEC 61300-3-26, *Fibre optic interconnecting devices and passive components – Basic test and measurement procedures – Part 3-26: Examinations and measurements – Measurement of the angular misalignment between fibre and ferrule axes*

IEC 61300-3-28, *Fibre optic interconnecting devices and passive components – Basic test and measurement procedures – Part 3-28: Examinations and measurements – Transient loss*

IEC 61300-3-34, *Fibre optic interconnecting devices and passive components – Basic test and measurement procedures – Part 3-34: Examinations and measurements – Attenuation of random mated connectors*

IEC 61300-3-35, *Fibre optic interconnecting devices and passive components – Basic test and measurement procedures – Part 3-35: Examinations and measurements – Visual inspection of fibre optic connectors and fibre-stub transceivers*

IEC 61300-3-47, *Fibre optic interconnecting devices and passive components – Basic test and measurement procedures – Part 3-47: Examinations and measurements – End face geometry of PC/APC spherically polished ferrules using interferometry*

IEC 61753-1, *Fibre optic interconnecting devices and passive components – Performance standard – Part 1: General and guidance for performance standards*

IEC 61753-021-2, *Fibre optic interconnecting devices and passive components – Performance standard – Part 021-2: Grade C/3 single-mode fibre optic connectors for category C – Controlled environment*

IEC 61754 (all parts), *Fibre optic interconnecting devices and passive components – Fibre optic connector interfaces*

IEC 61755 (all parts), *Fibre optic interconnecting devices and passive components – Connector optical interfaces*

IEC TR 61931, *Fibre optic – Terminology*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC TR 61931 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

3.1

terminated cable assembly

fibre optic cable terminated with any passive fibre optic component on each end

3.2

cord

cable terminated with fibre optic connectors at each end

EXAMPLE Equipment cord, work area cord or patchcord.

Note 1 to entry: Cord is also referred to as "terminated cable assembly".

[SOURCE: IEC 60794-2-51:2014, 3.2, modify – The definition has been rephrased, and an example and note to entry have been added.]

3.3

connector set

complete assembly of components (plug-adaptor-plug) required to provide demountable coupling between two or more optical fibres

3.4

reference cord

cord terminated with reference connector plugs

3.5

reference connector plug

connector plug manufactured with restricted tolerances for dimensions relevant to lateral and angular offset

Note 1 to entry: See IEC 61755-2-4 and IEC 61755-2-5.

3.6

change in attenuation

peak-to-peak variation

[SOURCE: IEC 61753-021-2:2007, 3.1]

4 Description

4.1 General

Patchcords, work area cords, equipment cords and reference cords (called "cords" in subsequent text) defined according to this document are terminated cable assemblies with optical connector plugs at each end.

The length, unless otherwise specified, is defined as being between the end faces of the connector plugs.

Cords, except reference cords, can be of any cable length. Reference cords have a length between 2 m and 5 m.

4.2 Optical fibres

Optical fibres meeting the requirements of IEC 60793-2-50 category B1.1 and B1.3 single-mode fibres shall be used. Once these cords are qualified, cords with the same construction using B6_a1 and B6_a2 fibre types according to IEC 60793-2-50 are qualified as well.

4.3 Cable design and construction

Cable used for the cords shall conform to the requirements of IEC 60794-2-50 and IEC 60794-2-51.

4.4 Optical connectors

4.4.1 Mechanical connectivity

The dimensional interface requirements in IEC 61754 (all parts) shall be met.

4.4.2 Optical performance requirements

The functionality of the connections according to this document is based upon physical contact. All the connector plugs shall conform to the standard performance grade as defined in IEC 61755 (all parts). Considered attenuation grades are R1 and R2 defined in IEC 61755-2-4 and IEC 61755-2-5, and B, C and D defined in IEC 61755-2-1 and IEC 61755-2-2. Considered return loss grades are 1, 2 and 3 defined in IEC 61755-2-1 and IEC 61755-2-2.

4.4.3 Connector set performance requirements

Connector sets shall conform to the requirements described in IEC 61753-021-2.

4.5 Cable bend radius

Care shall be taken to respect the minimum bend radius of the cable.

5 Tests

5.1 General

All tests and measurements have been selected from IEC 61300 (all parts) for connectors and from the cable test procedure outlined in IEC 60794-1-2. Additional requirements to certain tests are given in Annex C.

5.2 Measurement wavelengths

Unless otherwise specified in the individual test details, all attenuation measurements are made at the wavelengths given in Table 1.

Table 1 – Wavelengths for attenuation and return loss measurements

Fibre type	Centre wavelength		
	nm		
Single-mode	1 310	1 550	1 625

Return loss measurements shall be performed at the wavelengths specified in the individual tests.

5.3 Device under test

For this document, a device under test (DUT) is defined as a terminated cable assembly with optical connector plugs according to IEC 61754 (all parts) at all ends of the cord.

The sample size and product sourcing requirements are defined in Annex A.

The length of the DUT shall be 3,0 m to 5,0 m.

5.4 Test report

A fully documented test report and supporting data shall be prepared and shall be available for inspection as evidence that the tests described in this document have been carried out accordingly.

6 Test procedure

6.1 General

No deviation from the specified test method is allowed.

Unless otherwise specified, all tests shall be carried out at ambient temperature as specified in IEC 61300-1.

6.2 Visual examination

A visual examination shall be carried out on all DUTs before and after all mechanical and environmental tests (see Table 2). The outer cable sheath shall be marked at the end of the connector boot during the initial visual examination (see Annex B).

Table 2 – Visual examination requirements

No.	Test	Requirement	Details
1	Visual examination	No visible defects of cable or connector plugs	Method: IEC 61300-3-1 Examination: Product shall be visually checked without magnification

6.3 Fibre optic connector plug end face

The performance of the fibre optic connection depends on characteristics of the end faces of both connector plugs (see Table 3).

Table 3 – End face requirements

No.	Test	Requirement	Details
2	End face geometry	IEC 61755-3 (all parts)	Method: IEC 61300-3-47, End face geometry IEC 61300-3-25, Concentricity IEC 61300-3-26, Angular misalignment
3	Fibre optic connector end face visual inspection	IEC 61300-3-35	Method: IEC 61300-3-35 Examination: Scratches, defects, debris
4	Ferrule compression force ^a	IEC 61754 (all parts): for the connectorised buffered fibre IEC 60794-2-50: additional requirements for the ruggedised fibre cable	Method: IEC 61300-3-22 Examination: Movement length, compression force
^a This test is applicable to connector plugs with spring loaded ferrules.			

6.4 Optical performance requirements

Optical performance requirements for attenuation and return loss are given in the following Table 4. These requirements are related to connections between the same fibre types.

Table 4 – Optical performance requirements

No.	Test	Requirement	Details
5	Attenuation	R1 (reference grade): ≤ 0,1 dB R2 (reference grade): ≤ 0,2 dB Grade B: ≤ 0,12 dB mean ≤ 0,25 dB for 97 % Grade C: ≤ 0,25 dB mean ≤ 0,5 dB for 97 % Grade D: ≤ 0,5 dB mean ≤ 1,0 dB for 97 %	Method: IEC 61300-3-34, Method 2 Source type: LED/LD Wavelength: (1 310 ± 30) nm (1 550 ± 30) nm (1 625 ± 30) nm Source stability: ±0,01 dB over 1 h Detector linearity: ±0,01 dB over the dynamic range to be measured Launch fibre length: > 2 m. Only the fundamental mode shall propagate at the connector interface to be tested and at the detector Pre-conditioning procedure: Clean plug and adaptor according to manufacturer's instructions
6	Return loss	Grade 1: ≥ 60 dB Grade 2: ≥ 45 dB Grade 3: ≥ 35 dB	Method: IEC 61300-3-6, Method 1 Wavelengths: (1 310 ± 30) nm (1 550 ± 30) nm (1 625 ± 30) nm Source stability: ±0,01 dB over 1 h Detector linearity: ±0,1 dB over the dynamic range to be measured

6.5 Environmental performance requirements

Environmental performance requirements are given in the following Table 5.

Table 5 – Environmental performance requirements

No.	Test	Requirement	Details	
7	Change of temperature	<p>Change in attenuation during the test</p> <p>at (1 310 ± 30) nm ≤ 0,40 dB</p> <p>at (1 625 ± 30) nm ≤ 1,0 dB</p> <p>Change in attenuation before and after the test</p> <p>at (1 310 ± 30) nm ≤ 0,20 dB</p> <p>at (1 625 ± 30) nm ≤ 0,40 dB</p> <p>Initial and final attenuation shall be ≤ specified for the grade</p> <p>Return loss shall satisfy the requirements for the specified grade</p> <p>Final visual examination: see Annex B</p>	<p>Method:</p> <p>Low temperature:</p> <p>High temperature:</p> <p>Duration at temperature extreme:</p> <p>Rate of change of temperature:</p> <p>Number of cycles:</p> <p>DUT optically functioning:</p> <p>Measurements required:</p> <p>Sampling rate:</p> <p>Attenuation:</p> <p>Return loss:</p> <p>Pre-conditioning procedure:</p> <p>Recovery procedure:</p>	<p>IEC 61300-2-22, see Annex C</p> <p>–10 °C</p> <p>60 °C</p> <p>1 h</p> <p>1 °C/min</p> <p>5</p> <p>Yes</p> <p>Measuring procedure: IEC 61300-3-3. Measurements before, during and after the test</p> <p>Max. interval 10 min</p> <p>According to Table 4</p> <p>According to Table 4</p> <p>2 h at normal ambient conditions. Clean connector plugs and adaptor according to manufacturer's instructions</p> <p>2 h at normal ambient conditions. Connection shall not be demated</p>

6.6 Mechanical performance requirements

Mechanical performance requirements are given in the following Table 6.

Table 6 – Mechanical performance requirements

No.	Test	Requirement	Details	
8	Fibre/Cable retention	<p>Change in attenuation during the test</p> <p>at (1 310 ± 30) nm ≤ 0,20 dB</p> <p>at (1 625 ± 30) nm ≤ 0,50 dB</p> <p>Change in attenuation before and after the test</p> <p>at (1 310 ± 30) nm and (1 625 ± 30) nm ≤ 0,20 dB</p> <p>Initial and final attenuation shall be ≤ specified for the grade</p> <p>Return loss shall satisfy the requirements for the specified grade</p> <p>Final visual examination: see Annex B</p>	<p>Method:</p> <p>Tensile force:</p> <p>Point of application of the load:</p> <p>Duration of maximum load:</p> <p>DUT optically functioning:</p> <p>Measurements required:</p> <p>Sampling rate:</p> <p>Attenuation:</p> <p>Return loss:</p> <p>Pre-conditioning procedure:</p>	<p>IEC 61300-2-4</p> <p>50 N ± 2 N at 5 N/s</p> <p>0,3 m from connector plug</p> <p>The connector plug shall be rigidly mounted such that the load is applied to the fibre/cable retention mechanism and not to the coupling mechanism</p> <p>2 min at 50 N</p> <p>Yes</p> <p>Measuring procedure: IEC 61300-3-3. Measurements before, during and after the test</p> <p>Continuously</p> <p>According to Table 4</p> <p>According to Table 4</p> <p>Clean plug and adaptor according to manufacturer's instructions</p>
9	Static side load	<p>Change in attenuation during the test</p> <p>at (1 310 ± 30) nm ≤ 0,20 dB,</p> <p>at (1 625 ± 30) nm ≤ 0,50 dB</p> <p>Change in attenuation before and after the test</p> <p>at (1 310 ± 30) nm and (1 625 ± 30) nm ≤ 0,20 dB</p> <p>Initial and final attenuation shall be ≤ specified for the grade</p> <p>Return loss shall satisfy the requirements for the specified grade</p> <p>Final visual examination: see Annex B</p>	<p>Method:</p> <p>Magnitude of the load (90° to plug axis):</p> <p>Point of application of the load:</p> <p>Method of mounting:</p> <p>Duration of load:</p> <p>DUT optically functioning:</p> <p>Measurements required:</p> <p>Sampling rate:</p> <p>Attenuation:</p> <p>Return loss:</p> <p>Pre-conditioning procedure:</p>	<p>IEC 61300-2-42</p> <p>1 N</p> <p>0,2 m from rear of connector plug in two mutually perpendicular directions</p> <p>An adaptor shall be mounted rigidly to the mounting fixture</p> <p>1 h</p> <p>Yes</p> <p>Measuring procedure: IEC 61300-3-3. Measurements before, during and after the test</p> <p>Continuously</p> <p>According to Table 4</p> <p>According to Table 4</p> <p>Clean plug and adaptor according to manufacturer's instructions</p>

No.	Test	Requirement	Details	
10	Flexing strain relief of fibre optic devices	<p>Change in attenuation during the test</p> <p>at (1 310 ± 30) nm ≤ 0,20 dB,</p> <p>at (1 550 ± 30) nm ≤ 0,30 dB</p> <p>at (1 625 ± 30) nm ≤ 0,50 dB</p> <p>Change in attenuation before and after the test</p> <p>at (1 310 ± 30) nm, (1 550 ± 30) nm and (1 625 ± 30) nm ≤ 0,20 dB</p> <p>Initial and final attenuation shall be ≤ specified for the grade</p> <p>Return loss shall satisfy the requirements for the specified grade</p> <p>Final visual examination: see Annex B</p>	<p>Method:</p> <p>Magnitude of the load:</p> <p>Point of application of the load:</p> <p>Method of mounting:</p> <p>Cycle:</p> <p>Number of cycles:</p> <p>Cycling rate:</p> <p>DUT optically functioning:</p> <p>Measurements required:</p> <p>Sampling rate:</p> <p>Attenuation:</p> <p>Return loss:</p> <p>Pre-conditioning procedure:</p>	<p>IEC 61300-2-44</p> <p>2 N</p> <p>0,2 m from rear of connector plug</p> <p>An adaptor shall be mounted rigidly to the mounting fixture</p> <p>From 0° to -90° to 0° to +90° to 0°</p> <p>100</p> <p>20 cycles/min</p> <p>Yes</p> <p>Measuring procedure: IEC 61300-3-28. Measurements before, during and after the test</p> <p>Continuously</p> <p>According to Table 4</p> <p>According to Table 4</p> <p>Clean plug and adaptor according to manufacturer's instructions</p>

Annex A (normative)

Sample size requirements

Sample size for the cords shall be as indicated in the following Table A.1.

Table A.1 – Sample size requirements

No.	Test	Simplex	Duplex
1	Visual examination	15	10
2	End face geometry	15	10
3	Fibre optic cylindrical connector plug end face visual inspection	15	10
4	Ferrule compression force	15	10
5	Attenuation	15	10
6	Return loss	15	10
7	Change of temperature	6	3
8	Fibre/cable retention	6	3
10	Static side load	6	3
11	Flexing strain relief of fibre optic devices	6	3

The above tests are not intended to be performed in any particular sequence or grouping. They are intended to be performed individually; however, products from previous tests may be used if desired.

Annex B (normative)

Visual examination of outer cable sheath movement

B.1 Scope

This visual examination shall be made to ensure that the captivation or attachment of a cable sheath to a connector plug will withstand all environmental and mechanical tests required in this document.

B.2 Preparation of the DUT and initial visual examination

Preparation shall be made after initial visual examination (before all subsequent tests).

Mark the outer cable sheath at the end of the connector boot at both ends of the cable assembly as indicated in Figure B.1. The marks are required to identify movement of the cable sheath caused by environmental and mechanical stresses during subsequent tests.




Figure B.1 – Initial marking of the cable sheath

If the connector plug has a shrink-tube as part of the boot as, for example in some types of LC connector plugs, the mark shall be made on the outer cable sheath right at the end of the shrink-tube.

B.3 Final visual examination of outer cable sheath movement

Final visual examination shall be made after all tests have been finished. The outer sheath movement is visible through the movements of the marks at the outer cable sheath (see Figure B.2).

Requirements:

The allowed movement of the outer cable sheath relative to the connector boots (at least of its fixing point e.g. sheath crimp, shrink-tube or gluing) shall be 1 mm maximum at any connector plug.




Figure B.2 – Final visual examination

NOTE The SC connector in the Figures B.1 and B.2 is assumed as an example.

Annex C (normative)

Change of temperature

The additional requirements for the change of temperature test (see configuration in Figure C.1) shall be as follows.

- The whole length of the patchcord together with both connections shall be within the climatic chamber.
- The cable coils shall be free, without any cable reel, and supported horizontally in the climatic chamber. The winding radius shall be larger than 150 mm.




Figure C.1 – Change of temperature test configuration

Bibliography

IEC 60794-1-2: 2017, *Optical fibre cables – Part 1-2: Generic specification – Basic optical cable test procedures – General and guidance*

IEC 60794-2, *Optical fibre cables – Part 2: Indoor cables – Sectional specification*

IEC 61300 (all parts), *Fibre optic interconnecting devices and passive components – Basic test and measurement procedures*

IEC 61753 (all parts), *Fibre optic interconnecting devices and passive components – Performance standard*

IEC 61756-1, *Fibre optic interconnecting devices and passive components – Interface standard for fibre management systems – Part 1: General and guidance*

ISO/IEC 11801, *Information technology – Generic cabling for customer premises*
